

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122018\
 Data File : BE098532.D
 Acq On : 20 Dec 2018 10:46
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 20 13:21:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	48#	-0.01
2	1,4-Dioxane	0.692	0.766	-10.7	54	-0.01
3	n-Nitrosodimethylamine	0.623	0.603	3.2	50	-0.01
4 S	2-Fluorophenol	0.936	1.129	-20.6	60	0.00
5 S	Phenol-d6	1.326	1.644	-24.0	59	-0.01
6	bis(2-Chloroethyl)ether	1.260	1.194	5.2	47#	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	49#	-0.01
8 S	Nitrobenzene-d5	0.364	0.345	5.2	51	0.00
9	Naphthalene	4.025	4.201	-4.4	53	-0.01
10	Hexachlorobutadiene	0.256	0.279	-9.0	55	-0.01
11 SURR	2-Methylnaphthalene-d10	0.650	0.690	-6.2	53	-0.01
12	2-Methylnaphthalene	0.654	0.691	-5.7	53	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	54	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.190	-29.3#	74	-0.01
15 S	2-Fluorobiphenyl	1.687	1.574	6.7	56	-0.01
16	Acenaphthylene	1.874	1.739	7.2	52	-0.01
17	Acenaphthene	1.126	1.156	-2.7	56	-0.01
18	Fluorene	1.506	1.580	-4.9	57	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	56	-0.02
20	4-Bromophenyl-phenylether	0.215	0.213	0.9	58	-0.01
21	Hexachlorobenzene	0.231	0.240	-3.9	59	-0.01
22	Pentachlorophenol	0.046	0.078	-69.6#	121	-0.02
23	Phenanthrene	0.972	1.008	-3.7	59	-0.01
24	Anthracene	0.866	0.835	3.6	55	-0.01
25 SURR	Fluoranthene-d10	5.131	5.821	-13.4	61	-0.02
26	Fluoranthene	1.286	1.418	-10.3	59	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	72	-0.01
28	Pvrene	1.255	1.099	12.4	62	-0.01
29 S	Terphenyl-d14	0.972	0.883	9.2	66	-0.02
30	Benzo(a)anthracene	1.139	1.180	-3.6	76	-0.02
31	Chrysene	1.193	1.180	1.1	72	-0.01
32	Bis(2-ethylhexyl)phthalate	2.313	2.730	-18.0	86	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.009	15.0	65	-0.04
34 I	Perylene-d12	1.000	1.000	0.0	68	-0.03
35	Benzo(b)fluoranthene	1.094	1.175	-7.4	75	-0.03
36	Benzo(k)fluoranthene	1.274	1.426	-11.9	79	-0.03
37 C	Benzo(a)pyrene	1.129	1.116	1.2	69	-0.03
38	Dibenzo(a,h)anthracene	1.063	0.976	8.2	64	-0.04
39	Benzo(a,h,i)perylene	1.071	1.030	3.8	67	-0.05

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122018\
Data File : BE098532.D
Acq On : 20 Dec 2018 10:46
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 20 13:21:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			