

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122018\
 Data File : BE098545.D
 Acq On : 20 Dec 2018 18:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 21 05:03:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	61	-0.01
2	1,4-Dioxane	0.692	0.671	3.0	60	-0.01
3	n-Nitrosodimethylamine	0.623	0.494	20.7	52	0.00
4 S	2-Fluorophenol	0.936	0.923	1.4	62	0.00
5 S	Phenol-d6	1.326	1.574	-18.7	72	-0.01
6	bis(2-Chloroethyl)ether	1.260	1.101	12.6	55	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	67	-0.03
8 S	Nitrobenzene-d5	0.364	0.339	6.9	68	0.00
9	Naphthalene	4.025	4.034	-0.2	69	-0.01
10	Hexachlorobutadiene	0.256	0.271	-5.9	73	-0.01
11 SURR	2-Methylnaphthalene-d10	0.650	0.691	-6.3	72	-0.01
12	2-Methylnaphthalene	0.654	0.670	-2.4	70	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	78	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.181	-23.1	101	-0.01
15 S	2-Fluorobiphenyl	1.687	1.524	9.7	77	-0.01
16	Acenaphthylene	1.874	1.746	6.8	75	-0.01
17	Acenaphthene	1.126	1.133	-0.6	79	-0.01
18	Fluorene	1.506	1.568	-4.1	81	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	-0.02
20	4-Bromophenyl-phenylether	0.215	0.190	11.6	79	-0.01
21	Hexachlorobenzene	0.231	0.226	2.2	85	-0.01
22	Pentachlorophenol	0.046	0.065	-41.3#	154#	-0.02
23	Phenanthrene	0.972	0.979	-0.7	87	-0.01
24	Anthracene	0.866	0.807	6.8	80	-0.01
25 SURR	Fluoranthene-d10	5.131	5.885	-14.7	94	-0.02
26	Fluoranthene	1.286	1.462	-13.7	93	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	117	-0.01
28	Pvrene	1.255	1.077	14.2	98	-0.02
29 S	Terphenyl-d14	0.972	0.859	11.6	105	-0.02
30	Benzo(a)anthracene	1.139	1.139	0.0	120	-0.03
31	Chrysene	1.193	1.128	5.4	112	-0.01
32	Bis(2-ethylhexyl)phthalate	2.313	2.287	1.1	117	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	0.971	18.2	101	-0.04
34 I	Perylene-d12	1.000	1.000	0.0	110	-0.03
35	Benzo(b)fluoranthene	1.094	1.142	-4.4	119	-0.02
36	Benzo(k)fluoranthene	1.274	1.314	-3.1	118	-0.03
37 C	Benzo(a)pyrene	1.129	1.111	1.6	111	-0.03
38	Dibenzo(a,h)anthracene	1.063	0.945	11.1	101	-0.05
39	Benzo(a,h,i)perylene	1.071	0.961	10.3	102	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122018\
Data File : BE098545.D
Acq On : 20 Dec 2018 18:50
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 21 05:03:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			