

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122018\
 Data File : BE098545.D
 Acq On : 20 Dec 2018 18:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 21 05:03:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	61	-0.01
2	1,4-Dioxane	0.400	0.388	3.0	60	-0.01
3	n-Nitrosodimethylamine	0.400	0.317	20.8	52	0.00
4 S	2-Fluorophenol	0.400	0.394	1.5	62	0.00
5 S	Phenol-d6	0.400	0.475	-18.7	72	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.349	12.8	55	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	67	-0.03
8 S	Nitrobenzene-d5	0.400	0.373	6.8	68	0.00
9	Naphthalene	0.400	0.401	-0.3	69	-0.01
10	Hexachlorobutadiene	0.400	0.423	-5.7	73	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.425	-6.2	72	-0.01
12	2-Methylnaphthalene	0.400	0.410	-2.5	70	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	78	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.494	-23.5	101	-0.01
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	77	-0.01
16	Acenaphthylene	0.400	0.373	6.8	75	-0.01
17	Acenaphthene	0.400	0.403	-0.8	79	-0.01
18	Fluorene	0.400	0.416	-4.0	81	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	-0.02
20	4-Bromophenyl-phenylether	0.400	0.354	11.5	79	-0.01
21	Hexachlorobenzene	0.400	0.390	2.5	85	-0.01
22	Pentachlorophenol	0.400	0.647	-61.7#	154	-0.02
23	Phenanthrene	0.400	0.403	-0.8	87	-0.01
24	Anthracene	0.400	0.372	7.0	80	-0.01
25 SURR	Fluoranthene-d10	0.400	0.459	-14.8	94	-0.02
26	Fluoranthene	0.400	0.455	-13.7	93	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	117	-0.01
28	Pvrene	0.400	0.343	14.2	98	-0.02
29 S	Terphenyl-d14	0.400	0.353	11.8	105	-0.02
30	Benzo(a)anthracene	0.400	0.400	0.0	120	-0.03
31	Chrysene	0.400	0.378	5.5	112	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.395	1.3	117	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.327	18.3	101	-0.04
34 I	Perylene-d12	0.400	0.400	0.0	110	-0.03
35	Benzo(b)fluoranthene	0.400	0.418	-4.5	119	-0.02
36	Benzo(k)fluoranthene	0.400	0.412	-3.0	118	-0.03
37 C	Benzo(a)pyrene	0.400	0.394	1.5	111	-0.03
38	Dibenzo(a,h)anthracene	0.400	0.356	11.0	101	-0.05
39	Benzo(a,h,i)perylene	0.400	0.359	10.3	102	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122018\
Data File : BE098545.D
Acq On : 20 Dec 2018 18:50
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 21 05:03:45 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			