

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122118\
 Data File : BE098547.D
 Acq On : 21 Dec 2018 9:33
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 21 15:42:05 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	47#	-0.01
2	1,4-Dioxane	0.692	0.687	0.7	48#	-0.01
3	n-Nitrosodimethylamine	0.623	0.639	-2.6	52	0.00
4 S	2-Fluorophenol	0.936	1.004	-7.3	52	0.00
5 S	Phenol-d6	1.326	1.320	0.5	47#	-0.01
6	bis(2-Chloroethyl)ether	1.260	1.124	10.8	44#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	55	-0.01
8 S	Nitrobenzene-d5	0.364	0.309	15.1	51	0.00
9	Naphthalene	4.025	3.812	5.3	53	-0.01
10	Hexachlorobutadiene	0.256	0.261	-2.0	57	-0.01
11 SURR	2-Methylnaphthalene-d10	0.650	0.662	-1.8	56	-0.01
12	2-Methylnaphthalene	0.654	0.641	2.0	55	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	66	-0.01
14 S	2,4,6-Tribromophenol	0.147	0.159	-8.2	75	-0.01
15 S	2-Fluorobiphenyl	1.687	1.383	18.0	59	-0.01
16	Acenaphthylene	1.874	1.601	14.6	58	-0.01
17	Acenaphthene	1.126	1.031	8.4	61	-0.01
18	Fluorene	1.506	1.402	6.9	62	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	66	-0.01
20	4-Bromophenyl-phenylether	0.215	0.191	11.2	61	-0.01
21	Hexachlorobenzene	0.231	0.220	4.8	64	-0.01
22	Pentachlorophenol	0.046	0.056	-21.7	102	-0.01
23	Phenanthrene	0.972	0.882	9.3	61	-0.01
24	Anthracene	0.866	0.726	16.2	56	-0.01
25 SURR	Fluoranthene-d10	5.131	5.224	-1.8	64	-0.01
26	Fluoranthene	1.286	1.269	1.3	62	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	73	-0.01
28	Pvrene	1.255	1.114	11.2	63	-0.01
29 S	Terphenyl-d14	0.972	0.865	11.0	65	-0.01
30	Benzo(a)anthracene	1.139	0.992	12.9	65	-0.02
31	Chrysene	1.193	1.079	9.6	66	-0.01
32	Bis(2-ethylhexyl)phthalate	2.313	1.893	18.2	60	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	0.996	16.1	64	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	67	-0.02
35	Benzo(b)fluoranthene	1.094	1.001	8.5	64	-0.02
36	Benzo(k)fluoranthene	1.274	1.175	7.8	65	-0.02
37 C	Benzo(a)pyrene	1.129	1.024	9.3	62	-0.03
38	Dibenzo(a,h)anthracene	1.063	0.979	7.9	64	-0.04
39	Benzo(a,h,i)perylene	1.071	0.994	7.2	64	-0.04

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122118\
Data File : BE098547.D
Acq On : 21 Dec 2018 9:33
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 21 15:42:05 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			