

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122118\
 Data File : BE098557.D
 Acq On : 21 Dec 2018 18:24
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 22 00:58:01 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 10 14:59:55 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	52	-0.01
2	1,4-Dioxane	0.692	0.678	2.0	52	-0.01
3	n-Nitrosodimethylamine	0.623	0.431	30.8#	39#	0.01
4 S	2-Fluorophenol	0.936	0.900	3.8	52	0.00
5 S	Phenol-d6	1.326	1.299	2.0	51	-0.01
6	bis(2-Chloroethyl)ether	1.260	0.990	21.4	43#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	52	-0.01
8 S	Nitrobenzene-d5	0.364	0.299	17.9	47#	0.00
9	Naphthalene	4.025	3.830	4.8	51	-0.01
10	Hexachlorobutadiene	0.256	0.269	-5.1	56	-0.01
11 SURR	2-Methylnaphthalene-d10	0.650	0.624	4.0	51	-0.02
12	2-Methylnaphthalene	0.654	0.574	12.2	47#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	55	-0.02
14 S	2,4,6-Tribromophenol	0.147	0.167	-13.6	65	-0.01
15 S	2-Fluorobiphenyl	1.687	1.487	11.9	53	-0.02
16	Acenaphthylene	1.874	1.564	16.5	47#	0.00
17	Acenaphthene	1.126	0.997	11.5	49#	-0.02
18	Fluorene	1.506	1.347	10.6	49#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	51	-0.01
20	4-Bromophenyl-phenylether	0.215	0.191	11.2	48#	-0.01
21	Hexachlorobenzene	0.231	0.223	3.5	50	-0.01
22	Pentachlorophenol	0.046	0.055	-19.6	78	-0.01
23	Phenanthrene	0.972	0.890	8.4	48#	-0.01
24	Anthracene	0.866	0.708	18.2	42#	-0.01
25 SURR	Fluoranthene-d10	5.131	4.925	4.0	47#	-0.01
26	Fluoranthene	1.286	1.236	3.9	47#	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	63	-0.01
28	Pvrene	1.255	0.978	22.1	48#	-0.01
29 S	Terphenyl-d14	0.972	0.757	22.1	50#	-0.01
30	Benzo(a)anthracene	1.139	0.954	16.2	54	-0.01
31	Chrysene	1.193	1.063	10.9	57	-0.01
32	Bis(2-ethylhexyl)phthalate	2.313	1.587	31.4#	44#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.187	1.065	10.3	60	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	65	-0.02
35	Benzo(b)fluoranthene	1.094	1.005	8.1	62	-0.02
36	Benzo(k)fluoranthene	1.274	1.245	2.3	66	-0.02
37 C	Benzo(a)pyrene	1.129	1.005	11.0	59	-0.02
38	Dibenzo(a,h)anthracene	1.063	0.983	7.5	62	-0.03
39	Benzo(a,h,i)perylene	1.071	1.006	6.1	63	-0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122118\
Data File : BE098557.D
Acq On : 21 Dec 2018 18:24
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Dec 22 00:58:01 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE121018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 10 14:59:55 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			