


Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE122118\
 Data File : BE098546.D
 Acq On : 21 Dec 2018 8:19
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE121018.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 10 14:59:55 2018



AutoFind: Scans 2574, 2575, 2576; Background Corrected with Scan 2567

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.8	7249	PASS
68	69	0.00	2	1.6	106	PASS
69	198	0.00	100	47.5	6517	PASS
70	69	0.00	2	1.3	82	PASS
127	198	10	80	48.3	6628	PASS
197	198	0.00	2	0.5	74	PASS
198	198	100	100	100.0	13720	PASS
199	198	5	9	7.2	981	PASS
275	198	10	60	29.2	4006	PASS
365	198	1	100	5.0	687	PASS
441	198	0.01	100	15.1	2068	PASS
442	198	50	100	84.0	11525	PASS
443	442	15	24	19.9	2293	PASS