

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122818\
 Data File : BE098570.D
 Acq On : 28 Dec 2018 12:27
 Operator : SJ/JU
 Sample : SSTD3.207
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD3.207

Quant Time: Dec 28 13:03:53 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE122818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 28 11:34:21 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.81	152	930	0.40	ng/ul	0.00
2) Naphthalene-d8	10.61	136	4223	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.47	164	2702	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.23	188	6549	0.40	ng/ul	0.00
16) Chrysene-d12	21.41	240	8117	0.40	ng/ul	0.00
20) Perylene-d12	23.92	264	7455	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng/ul	
14) Fluoranthene-d10	0.00	212	0d	0.00	ng/ul	
Target Compounds						
11) Pentachlorophenol	16.88	266	4577	13.358	ng/ul	Ovalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE122818\
Data File : BE098570.D
Acq On : 28 Dec 2018 12:27
Operator : SJ/JU
Sample : SSTD3.207
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD3.207

Quant Time: Dec 28 13:03:53 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE122818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 28 11:34:21 2018
Response via : Initial Calibration

