

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011720\
 Data File : BE101086.D
 Acq On : 17 Jan 2020 15:53
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 17 23:25:50 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	48#	0.00
2	1,4-Dioxane	1.020	1.050	-2.9	54	0.00
3	n-Nitrosodimethylamine	0.556	0.591	-6.3	59	0.00
4 S	2-Fluorophenol	0.900	1.136	-26.2#	61	0.00
5 S	Phenol-d6	1.139	1.132	0.6	54	0.00
6	bis(2-Chloroethyl)ether	1.357	1.303	4.0	51	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	52	0.00
8 S	Nitrobenzene-d5	0.286	0.260	9.1	53	0.00
9	Naphthalene	4.441	4.469	-0.6	51	0.00
10	Hexachlorobutadiene	0.188	0.196	-4.3	52	-0.01
11 SURR	2-Methylnaphthalene-d10	0.763	0.774	-1.4	52	-0.01
12	2-Methylnaphthalene	0.672	0.674	-0.3	53	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	67	0.00
14 S	2,4,6-Tribromophenol	0.113	0.105	7.1	71	0.00
15 S	2-Fluorobiphenyl	1.519	1.308	13.9	57	-0.01
16	Acenaphthylene	1.629	1.451	10.9	66	-0.01
17	Acenaphthene	1.166	1.091	6.4	64	-0.01
18	Fluorene	1.474	1.482	-0.5	68	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.195	0.153	21.5	72	-0.01
21	Hexachlorobenzene	0.214	0.183	14.5	74	-0.01
22	Pentachlorophenol	0.053	0.025	52.8#	54	-0.01
23	Phenanthrene	1.092	0.987	9.6	80	0.00
24	Anthracene	1.031	0.998	3.2	89	0.00
25 SURR	Fluoranthene-d10	5.555	6.448	-16.1	105	-0.01
26	Fluoranthene	1.368	1.634	-19.4	111	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	146	-0.01
28	Pvrene	1.489	1.201	19.3	115	0.00
29 S	Terphenyl-d14	0.982	0.887	9.7	130	-0.01
30	Benzo(a)anthracene	1.141	1.168	-2.4	155#	0.00
31	Chrysene	1.696	1.712	-0.9	143	-0.01
32	Bis(2-ethylhexyl)phthalate	1.902	2.217	-16.6	195#	0.00
33	Indeno(1,2,3-cd)pyrene	1.340	0.966	27.9#	108	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	119	-0.01
35	Benzo(b)fluoranthene	1.122	1.071	4.5	122	-0.01
36	Benzo(k)fluoranthene	1.626	1.456	10.5	107	-0.01
37 C	Benzo(a)pyrene	1.085	1.079	0.6	122	-0.01
38	Dibenzo(a,h)anthracene	0.998	0.826	17.2	107	-0.02
39	Benzo(a,h,i)perylene	1.227	0.979	20.2	93	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011720\
Data File : BE101086.D
Acq On : 17 Jan 2020 15:53
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jan 17 23:25:50 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			