


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE011720\
 Data File : BE101085.D
 Acq On : 17 Jan 2020 15:12
 Operator : JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE010820.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 08 18:36:25 2020



AutoFind: Scans 2546, 2547, 2548; Background Corrected with Scan 2538

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.2	32073	PASS
68	69	0.00	2	1.7	561	PASS
69	198	0.00	100	61.1	33643	PASS
70	69	0.00	2	0.7	240	PASS
127	198	10	80	58.8	32405	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	55101	PASS
199	198	5	9	6.9	3805	PASS
275	198	10	60	24.9	13741	PASS
365	198	1	100	4.0	2225	PASS
441	198	0.01	100	14.4	7925	PASS
442	198	50	100	89.7	49409	PASS
443	442	15	24	19.3	9545	PASS