

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE020720\
 Data File : BE101142.D
 Acq On : 7 Feb 2020 18:33
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 08 01:09:03 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 08 18:36:25 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	38	-0.03
2	1,4-Dioxane	0.400	0.507	-26.7#	46	-0.03
3	n-Nitrosodimethylamine	0.400	0.349	12.8	39	-0.02
4 S	2-Fluorophenol	0.400	0.385	3.8	37	-0.02
5 S	Phenol-d6	0.400	0.327	18.3	35	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.559	-39.8#	59	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	41	-0.01
8 S	Nitrobenzene-d5	0.400	0.354	11.5	40	-0.01
9	Naphthalene	0.400	0.391	2.3	39	-0.03
10	Hexachlorobutadiene	0.400	0.440	-10.0	43	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	39	-0.03
12	2-Methylnaphthalene	0.400	0.376	6.0	39	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	43	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.244	39.0#	31	-0.01
15 S	2-Fluorobiphenyl	0.400	0.378	5.5	40	-0.03
16	Acenaphthylene	0.400	0.388	3.0	47	-0.02
17	Acenaphthene	0.400	0.390	2.5	43	-0.03
18	Fluorene	0.400	0.393	1.8	43	-0.03
19 I	Phenanthrene-d10	0.400	0.400	0.0	66	-0.01
20	4-Bromophenyl-phenylether	0.400	0.340	15.0	59	-0.03
21	Hexachlorobenzene	0.400	0.286	28.5#	47	-0.03
22	Pentachlorophenol	0.400	0.253	36.8#	1	-0.05
23	Phenanthrene	0.400	0.344	14.0	58	-0.01
24	Anthracene	0.400	0.396	1.0	69	-0.01
25 SURR	Fluoranthene-d10	0.400	0.362	9.5	62	-0.02
26	Fluoranthene	0.400	0.357	10.8	63	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	71	-0.01
28	Pvrene	0.400	0.374	6.5	65	-0.02
29 S	Terphenyl-d14	0.400	0.420	-5.0	74	-0.02
30	Benzo(a)anthracene	0.400	0.356	11.0	66	-0.01
31	Chrysene	0.400	0.413	-3.2	71	-0.03
32	Bis(2-ethylhexyl)phthalate	0.400	0.532	-33.0#	109	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.471	-17.7	86	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	87	-0.02
35	Benzo(b)fluoranthene	0.400	0.332	17.0	77	-0.02
36	Benzo(k)fluoranthene	0.400	0.338	15.5	73	-0.02
37 C	Benzo(a)pyrene	0.400	0.397	0.8	89	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.391	2.3	92	-0.02
39	Benzo(a,h,i)perylene	0.400	0.360	10.0	76	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE020720\
Data File : BE101142.D
Acq On : 7 Feb 2020 18:33
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 08 01:09:03 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE010820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 08 18:36:25 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				