


Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE020720\
 Data File : BE101134.D
 Acq On : 7 Feb 2020 12:23
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE010820.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jan 08 18:36:25 2020



AutoFind: Scans 2544, 2545, 2546; Background Corrected with Scan 2536

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	53.9	13639	PASS
68	69	0.00	2	1.7	238	PASS
69	198	0.00	100	56.5	14299	PASS
70	69	0.00	2	0.6	81	PASS
127	198	10	80	56.2	14211	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	25303	PASS
199	198	5	9	6.7	1690	PASS
275	198	10	60	25.3	6406	PASS
365	198	1	100	3.7	934	PASS
441	198	0.01	100	14.6	3700	PASS
442	198	50	100	89.7	22705	PASS
443	442	15	24	18.9	4288	PASS