

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032421\
 Data File : BE101151.D
 Acq On : 24 Mar 2021 17:26
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Mar 24 18:10:34 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 16:00:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.850	152	4124	0.40	ng	0.00
7) Naphthalene-d8	10.641	136	17099	0.40	ng	0.00
13) Acenaphthene-d10	14.472	164	10870	0.40	ng	0.00
19) Phenanthrene-d10	17.225	188	28627	0.40	ng	0.00
29) Chrysene-d12	21.378	240	30446	0.40	ng	0.00
36) Perylene-d12	23.863	264	28796	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.431	112	4489	0.46	ng	0.00
5) Phenol-d6	6.997	99	6887	0.46	ng	0.00
8) Nitrobenzene-d5	8.991	82	4912	0.48	ng	0.00
11) 2-Methylnaphthalene-d10	12.210	152	11893	0.43	ng	0.00
14) 2,4,6-Tribromophenol	15.970	330	640	0.33	ng	0.00
15) 2-Fluorobiphenyl	13.103	172	17565	0.43	ng	0.00
27) Fluoranthene-d10	19.232	212	150733	0.42	ng	0.00
31) Terphenyl-d14	19.835	244	32408	0.47	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.381	88	3085	0.46	ng	97
3) n-Nitrosodimethylamine	3.681	42	3432	0.49	ng	# 94
6) bis(2-Chloroethyl)ether	7.274	93	5744	0.45	ng	98
9) Naphthalene	10.680	128	78341	0.45	ng	100
10) Hexachlorobutadiene	10.979	225	3487	0.44	ng	100
12) 2-Methylnaphthalene	12.282	142	13043	0.44	ng	99
16) Acenaphthylene	14.184	152	16755	0.40	ng	100
17) Acenaphthene	14.530	154	12390	0.41	ng	100
18) Fluorene	15.532	166	16138	0.41	ng	100
20) 4,6-Dinitro-2-methylph...	15.592	198	680	0.47	ng	# 85
21) 4-Bromophenyl-phenylether	16.424	248	4463	0.35	ng	98
22) Hexachlorobenzene	16.530	284	5805	0.41	ng	98
23) Atrazine	16.681	200	3247	0.40	ng	98
24) Pentachlorophenol	16.877	266	1233	0.50	ng	96
25) Phenanthrene	17.255	178	34462	0.43	ng	99
26) Anthracene	17.346	178	28857	0.42	ng	100
28) Fluoranthene	19.261	202	42361	0.42	ng	100
30) Pyrene	19.623	202	42151	0.47	ng	100
32) Benzo(a)anthracene	21.354	228	35913	0.46	ng	99
33) Chrysene	21.415	228	42128	0.45	ng	100
34) Bis(2-ethylhexyl)phtha...	21.306	149	21176	0.41	ng	100
35) Indeno(1,2,3-cd)pyrene	26.487	276	28604	0.47	ng	100
37) Benzo(b)fluoranthene	23.100	252	30462	0.40	ng	99
38) Benzo(k)fluoranthene	23.150	252	40878	0.42	ng	99
39) Benzo(a)pyrene	23.749	252	26659	0.38	ng	99
40) Dibenzo(a,h)anthracene	26.520	278	23389	0.42	ng	99
41) Benzo(g,h,i)perylene	27.293	276	30483	0.42	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032421\
Data File : BE101151.D
Acq On : 24 Mar 2021 17:26
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Mar 24 18:10:34 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Mar 24 16:00:06 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.850 min Scan# 433
 Delta R.T. -0.001 min
 Lab File: BE101151.D
 Acq: 24 Mar 2021 17:26




Instrument :
 BNA_E
 ClientSampleId :


Tgt Ion:152 Resp: 4124
 Ion Ratio Lower Upper
 152 100
 150 155.5 126.3 189.5
 115 60.5 48.9 73.3



1,4-Dioxane
 Concen: 0.46 ng
 RT: 3.381 min Scan# 45
 Delta R.T. -0.002 min
 Lab File: BE101151.D
 Acq: 24 Mar 2021 17:26

Tgt Ion: 88 Resp: 3085
 Ion Ratio Lower Upper
 88 100
 58 89.0 68.5 102.7
 43 35.5 27.2 40.8









n-Nitrosodimethylamine
Concen: 0.49 ng
RT: 3.681 min Scan# 71
Delta R.T. -0.002 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26




Instrument :
BNA_E
ClientSampleId :


Tgt Ion:	42	Resp:	3432
Ion	Ratio	Lower	Upper
42	100		
74	102.8	87.0	130.6
44	14.7	8.3	12.5#



2-Fluorophenol
Concen: 0.46 ng
RT: 5.431 min Scan# 223
Delta R.T. -0.002 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Tgt Ion:	112	Resp:	4489
Ion	Ratio	Lower	Upper
112	100		
64	74.9	60.8	91.2
63	151.1	119.4	179.0









Phenol-d6
 Concen: 0.46 ng
 RT: 6.997 min Scan# 359
 Delta R.T. -0.002 min
 Lab File: BE101151.D
 Acq: 24 Mar 2021 17:26




Instrument :
 BNA_E
 ClientSampleId :

Tgt Ion:	99	Resp:	6887
Ion	Ratio	Lower	Upper
99	100		
42	24.5	18.8	28.2
71	33.1	26.3	39.5




bis(2-Chloroethyl)ether
 Concen: 0.45 ng
 RT: 7.274 min Scan# 383
 Delta R.T. -0.002 min
 Lab File: BE101151.D
 Acq: 24 Mar 2021 17:26

Tgt Ion:	93	Resp:	5744
Ion	Ratio	Lower	Upper
93	100		
63	370.5	292.5	438.7
95	32.9	25.8	38.6





Abundance Scan 652 (10.640 min): BE101146.D\data.ms (-6447-)




Naphthalene-d8
Concen: 0.40 ng
RT: 10.641 min Scan# 652
Delta R.T. 0.001 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampleId :


Tgt Ion:	136	Resp:	17099
Ion	Ratio	Lower	Upper
136	100		
137	11.8	8.9	13.3
54	24.0	20.2	30.4
68	5.7	4.0	6.0



Abundance Scan 652 (10.641 min): BE101151.D\data.ms (-637-)





Abundance Scan 525 (8.991 min): BE101146.D\data.ms (-5227-)




Nitrobenzene-d5
Concen: 0.48 ng
RT: 8.991 min Scan# 525
Delta R.T. 0.001 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26


Tgt Ion:	82	Resp:	4912
Ion	Ratio	Lower	Upper
82	100		
128	130.8	101.5	152.3
54	221.1	189.8	284.6



Abundance Scan 525 (8.991 min): BE101151.D\data.ms (-510-)





Abundance Scan 655 (10.679 min): BE101146.D\data.ms (-650) (-)




Naphthalene
Concen: 0.45 ng
RT: 10.680 min Scan# 655
Delta R.T. 0.001 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampled :


Tgt Ion:	128	Resp:	78341
Ion	Ratio	Lower	Upper
128	100		
129	2.9	2.2	3.2
127	3.3	2.7	4.1



Abundance Scan 655 (10.680 min): BE101151.D\data.ms (-647) (-)





Abundance Scan 678 (10.978 min): BE101146.D\data.ms (-673) (-)





Hexachlorobutadiene
Concen: 0.44 ng
RT: 10.979 min Scan# 678
Delta R.T. 0.001 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Tgt Ion:	225	Resp:	3487
Ion	Ratio	Lower	Upper
225	100		
223	62.8	50.1	75.1
227	64.7	51.7	77.5



Abundance Scan 678 (10.979 min): BE101151.D\data.ms (-670) (-)









2-Methylnaphthalene-d10
Concen: 0.43 ng
RT: 12.210 min Scan# 764
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26




Instrument :
BNA_E
ClientSampled :


Tgt Ion:152 Resp: 11893
Ion Ratio Lower Upper
152 100
151 20.8 16.6 24.8



2-Methylnaphthalene
Concen: 0.44 ng
RT: 12.282 min Scan# 769
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Tgt Ion:142 Resp: 13043
Ion Ratio Lower Upper
142 100
141 90.5 71.8 107.6
115 34.3 27.8 41.6









Acenaphthene-d10
Concen: 0.40 ng
RT: 14.472 min Scan# 921
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26




Instrument :
BNA_E
ClientSampleId :


Tgt Ion:	164	Resp:	10870
Ion	Ratio	Lower	Upper
164	100		
162	98.5	76.6	115.0
160	43.3	34.5	51.7



2,4,6-Tribromophenol
Concen: 0.33 ng
RT: 15.970 min Scan# 1017
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Tgt Ion:	330	Resp:	640
Ion	Ratio	Lower	Upper
330	100		
332	99.8	69.1	103.7
141	41.7	30.6	45.8









2-Fluorobiphenyl
Concen: 0.43 ng
RT: 13.103 min Scan# 826
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26




Instrument :
BNA_E
ClientSampled :


Tgt Ion:	172	Resp:	17565
Ion	Ratio	Lower	Upper
172	100		
171	33.2	26.4	39.6
170	21.8	17.4	26.0



Acenaphthylene
Concen: 0.40 ng
RT: 14.184 min Scan# 901
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Tgt Ion:	152	Resp:	16755
Ion	Ratio	Lower	Upper
152	100		
151	19.1	15.4	23.0
153	13.1	10.3	15.5









Acenaphthene
Concen: 0.41 ng
RT: 14.530 min Scan# 925
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26




Instrument :
BNA_E
ClientSampleId :


Tgt Ion:	154	Resp:	12390
Ion	Ratio	Lower	Upper
154	100		
153	108.2	86.7	130.1
152	56.9	46.1	69.1



Fluorene
Concen: 0.41 ng
RT: 15.532 min Scan# 988
Delta R.T. 0.015 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Tgt Ion:	166	Resp:	16138
Ion	Ratio	Lower	Upper
166	100		
165	98.8	78.9	118.3
167	13.4	10.7	16.1









Phenanthrene-d10
 Concen: 0.40 ng
 RT: 17.225 min Scan# 1100
 Delta R.T. -0.000 min
 Lab File: BE101151.D
 Acq: 24 Mar 2021 17:26




Instrument :
 BNA_E
 ClientSampleId :

Tgt Ion:	188	Resp:	28627
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	6.1	4.8	7.2



4,6-Dinitro-2-methylphenol
 Concen: 0.47 ng
 RT: 15.592 min Scan# 992
 Delta R.T. -0.000 min
 Lab File: BE101151.D
 Acq: 24 Mar 2021 17:26


Tgt Ion:	198	Resp:	680
Ion	Ratio	Lower	Upper
198	100		
51	65.9	68.8	103.2#
105	40.2	32.7	49.1






Abundance Scan 1047 (16.424 min): BE101146.D\data.ms (-1024) (-)

4-Bromophenyl-phenylether
Concen: 0.35 ng
RT: 16.424 min Scan# 1047
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampleId :




Tgt Ion:	248	Resp:	4463
Ion	Ratio	Lower	Upper
248	100		
250	98.7	79.0	118.6
141	35.8	24.9	37.3






Abundance Scan 1054 (16.530 min): BE101146.D\data.ms (-1024) (-)


Hexachlorobenzene
Concen: 0.41 ng
RT: 16.530 min Scan# 1054
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26



Tgt Ion:	284	Resp:	5805
Ion	Ratio	Lower	Upper
284	100		
142	33.7	28.3	42.5
249	35.0	27.1	40.7





Abundance Scan 1064 (16.681 min): BE101146.D\data.ms (-1031) (-)




Atrazine
Concen: 0.40 ng
RT: 16.681 min Scan# 1064
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampleId :


Tgt Ion:	200	Resp:	3247
Ion Ratio	Lower	Upper	
200	100		
173	31.8	23.9	35.9
215	51.5	40.6	61.0



Abundance Scan 1064 (16.681 min): BE101151.D\data.ms (-1031) (-)





Abundance Scan 1076 (16.863 min): BE101146.D\data.ms (-1021) (-)




Pentachlorophenol
Concen: 0.50 ng
RT: 16.877 min Scan# 1077
Delta R.T. 0.014 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26


Tgt Ion:	266	Resp:	1233
Ion Ratio	Lower	Upper	
266	100		
264	60.8	51.4	77.0
268	64.6	49.9	74.9



Abundance Scan 1077 (16.877 min): BE101151.D\data.ms (-1069) (-)





Abundance Scan 1102 (17.256 min): BE101146.D\data.ms (-1082) (-)




Phenanthrene
Concen: 0.43 ng
RT: 17.255 min Scan# 1102
Delta R.T. -0.001 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampleId :


Tgt Ion:	178	Resp:	34462
Ion	Ratio	Lower	Upper
178	100		
176	18.7	14.7	22.1
179	15.0	12.4	18.6



Abundance Scan 1102 (17.255 min): BE101151.D\data.ms (-1082) (-)





Abundance Scan 1108 (17.346 min): BE101146.D\data.ms (-1101) (-)




Anthracene
Concen: 0.42 ng
RT: 17.346 min Scan# 1108
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26


Tgt Ion:	178	Resp:	28857
Ion	Ratio	Lower	Upper
178	100		
176	18.2	14.4	21.6
179	15.3	12.2	18.4



Abundance Scan 1108 (17.346 min): BE101151.D\data.ms (-1101) (-)






Abundance Scan 1361 (19.232 min): BE101146.D\data.ms (-1325) (-)




Fluoranthene-d10
Concen: 0.42 ng
RT: 19.232 min Scan# 1361
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampleId :

Tgt Ion:	212	Resp:	150733
Ion	Ratio	Lower	Upper
212	100		
106	3.1	2.6	3.8
104	1.8	1.4	2.2






Abundance Scan 1366 (19.261 min): BE101146.D\data.ms (-1329) (-)



Fluoranthene
Concen: 0.42 ng
RT: 19.261 min Scan# 1366
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26


Tgt Ion:	202	Resp:	42361
Ion	Ratio	Lower	Upper
202	100		
101	11.0	8.9	13.3
203	17.3	13.8	20.6






Abundance Scan 1605 (21.379 min): BE101146.D\data.ms (-1529) (-)

Chrysene-d12
Concen: 0.40 ng
RT: 21.378 min Scan# 1605
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampleId :




Tgt Ion:	240	Resp:	30446
Ion Ratio	Lower	Upper	
240	100		
120	5.1	4.1	6.1
236	25.9	20.3	30.5






Abundance Scan 1429 (19.622 min): BE101146.D\data.ms (-1439) (-)

Pyrene
Concen: 0.47 ng
RT: 19.623 min Scan# 1429
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26




Tgt Ion:	202	Resp:	42151
Ion Ratio	Lower	Upper	
202	100		
200	20.4	16.2	24.2
203	17.5	13.9	20.9






Abundance Scan 1466 (19.835 min): BE101146.D\data.ms (-1438) (-)

Terphenyl-d14
Concen: 0.47 ng
RT: 19.835 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampled :







Tgt Ion	Ratio	Lower	Upper
244	100		
212	26.5	21.7	32.5
122	8.6	6.6	10.0




Abundance Scan 1603 (21.354 min): BE101146.D\data.ms (-1594) (-)

Benzo(a)anthracene
Concen: 0.46 ng
RT: 21.354 min Scan# 1603
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26


Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.8	21.8	32.8
229	19.4	15.4	23.0








Scan 1608 (21.415 min): BE101146.D\data.ms (-1600) (-)
 Chrysene
 Concen: 0.45 ng
 RT: 21.415 min Scan# 1608
 Delta R.T. -0.000 min
 Lab File: BE101151.D
 Acq: 24 Mar 2021 17:26

Instrument :
 BNA_E
 ClientSampled :






Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.5	22.8	34.2
229	20.1	16.2	24.2




Scan 1599 (21.306 min): BE101146.D\data.ms (-1599) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 0.41 ng
 RT: 21.306 min Scan# 1599
 Delta R.T. -0.000 min
 Lab File: BE101151.D
 Acq: 24 Mar 2021 17:26

Tgt Ion	Ratio	Lower	Upper
149	100		
167	6.9	5.4	8.2
279	1.2	0.9	1.3




Abundance Scan 2903 (26.484 min): BE101146.D\data.ms (-2875) (-)





Indeno(1,2,3-cd)pyrene
Concen: 0.47 ng
RT: 26.487 min Scan# 2904
Delta R.T. 0.004 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26


Instrument :
BNA_E
ClientSampleId :




Tgt Ion	Ratio	Lower	Upper
276	100		
138	21.3	17.1	25.7
277	24.5	19.8	29.8





Abundance Scan 2172 (23.863 min): BE101146.D\data.ms (-2130) (-)




Perylene-d12
Concen: 0.40 ng
RT: 23.863 min Scan# 2172
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26



Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.9	20.5	30.7
265	25.8	20.6	30.8






Abundance Scan 1958 (23.100 min): BE101146.D\data.ms (-1937)




Benzo(b)fluoranthene
Concen: 0.40 ng
RT: 23.100 min Scan# 1958
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	17.8	26.8
125	9.3	7.5	11.3

Instrument :
BNA_E
ClientSampleId :






Abundance Scan 1972 (23.150 min): BE101146.D\data.ms (-1938)



Benzo(k)fluoranthene
Concen: 0.42 ng
RT: 23.150 min Scan# 1972
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26


Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	18.2	27.4
125	9.6	7.5	11.3






Abundance Scan 2140 (23.749 min): BE101146.D\data.ms (-289) (-)

Benzo(a)pyrene
Concen: 0.38 ng
RT: 23.749 min Scan# 2140
Delta R.T. -0.000 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampleId :




Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.6	18.6	28.0
125	10.9	8.3	12.5






Abundance Scan 2912 (26.516 min): BE101146.D\data.ms (-289) (-)


Dibenzo(a,h)anthracene
Concen: 0.42 ng
RT: 26.520 min Scan# 2913
Delta R.T. 0.004 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26



Tgt Ion	Ratio	Lower	Upper
278	100		
139	15.5	12.0	18.0
279	24.9	20.2	30.4




Abundance Scan 3129 (27.290 min): BE101146.D\data.ms (-3104)



Benzo(g,h,i)perylene
Concen: 0.42 ng
RT: 27.293 min Scan# 3130
Delta R.T. 0.004 min
Lab File: BE101151.D
Acq: 24 Mar 2021 17:26

Instrument :
BNA_E
ClientSampleId :



Tgt Ion:	276	Resp:	30483
Ion	Ratio	Lower	Upper
276	100		
277	25.6	19.7	29.5
138	19.3	15.3	22.9

