

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032421\
 Data File : BE101151.D
 Acq On : 24 Mar 2021 17:26
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE032421

Quant Time: Mar 24 18:12:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 16:00:06 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.649	0.748	-15.3	103	0.00
3	n-Nitrosodimethylamine	0.677	0.832	-22.9	115	0.00
4 S	2-Fluorophenol	0.943	1.089	-15.5	103	0.00
5 S	Phenol-d6	1.466	1.670	-13.9	103	0.00
6	bis(2-Chloroethyl)ether	1.226	1.393	-13.6	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.239	0.287	-20.1	125	0.00
9	Naphthalene	4.073	4.582	-12.5	103	0.00
10	Hexachlorobutadiene	0.185	0.204	-10.3	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.653	0.696	-6.6	100	0.00
12	2-Methylnaphthalene	0.701	0.763	-8.8	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.071	0.059	16.9	95	0.00
15 S	2-Fluorobiphenyl	1.501	1.616	-7.7	104	0.00
16	Acenaphthylene	1.539	1.541	-0.1	102	0.00
17	Acenaphthene	1.106	1.140	-3.1	105	0.00
18	Fluorene	1.445	1.485	-2.8	101	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	111	0.00
20	4,6-Dinitro-2-methylphenol	0.020	0.024	-20.0	126	0.00
21	4-Bromophenyl-phenylether	0.180	0.156	13.3	91	0.00
22	Hexachlorobenzene	0.196	0.203	-3.6	105	0.00
23	Atrazine	0.112	0.113	-0.9	113	0.00
24	Pentachlorophenol	0.052	0.043	17.3	111	0.01
25	Phenanthrene	1.114	1.204	-8.1	110	0.00
26	Anthracene	0.951	1.008	-6.0	113	0.00
27 SURR	Fluoranthene-d10	5.053	5.265	-4.2	106	0.00
28	Fluoranthene	1.414	1.480	-4.7	106	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
30	Pyrene	1.181	1.384	-17.2	105	0.00
31 S	Terphenyl-d14	0.909	1.064	-17.1	105	0.00
32	Benzo(a)anthracene	1.022	1.180	-15.5	113	0.00
33	Chrysene	1.225	1.384	-13.0	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.676	0.696	-3.0	111	0.00
35	Indeno(1,2,3-cd)pyrene	0.807	0.939	-16.4	116	0.00
36 I	Perylene-d12	1.000	1.000	0.0	112	0.00
37	Benzo(b)fluoranthene	1.049	1.058	-0.9	109	0.00
38	Benzo(k)fluoranthene	1.343	1.420	-5.7	110	0.00
39 C	Benzo(a)pyrene	0.983	0.926	5.8	104	0.00
40	Dibenzo(a,h)anthracene	0.804	0.812	-1.0	116	0.00
41	Benzo(g,h,i)perylene	1.009	1.059	-5.0	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

ClientSampleId :

ICVBE032421