

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032421\
 Data File : BE101151.D
 Acq On : 24 Mar 2021 17:26
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE032421

Quant Time: Mar 24 18:12:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE032421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 24 16:00:06 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.461	-15.3	103	0.00
3	n-Nitrosodimethylamine	0.400	0.491	-22.7	115	0.00
4 S	2-Fluorophenol	0.400	0.462	-15.5	103	0.00
5 S	Phenol-d6	0.400	0.456	-14.0	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.455	-13.7	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.481	-20.2	125	0.00
9	Naphthalene	0.400	0.450	-12.5	103	0.00
10	Hexachlorobutadiene	0.400	0.441	-10.2	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.426	-6.5	100	0.00
12	2-Methylnaphthalene	0.400	0.436	-9.0	104	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	106	0.00
14 S	2,4,6-Tribromophenol	0.400	0.330	17.5	95	0.00
15 S	2-Fluorobiphenyl	0.400	0.431	-7.7	104	0.00
16	Acenaphthylene	0.400	0.401	-0.3	102	0.00
17	Acenaphthene	0.400	0.412	-3.0	105	0.00
18	Fluorene	0.400	0.411	-2.7	101	0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	111	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.468	-17.0	126	0.00
21	4-Bromophenyl-phenylether	0.400	0.347	13.3	91	0.00
22	Hexachlorobenzene	0.400	0.414	-3.5	105	0.00
23	Atrazine	0.400	0.405	-1.3	113	0.00
24	Pentachlorophenol	0.400	0.499	-24.7	111	0.01
25	Phenanthrene	0.400	0.432	-8.0	110	0.00
26	Anthracene	0.400	0.424	-6.0	113	0.00
27 SURR	Fluoranthene-d10	0.400	0.417	-4.2	106	0.00
28	Fluoranthene	0.400	0.419	-4.7	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	102	0.00
30	Pyrene	0.400	0.469	-17.2	105	0.00
31 S	Terphenyl-d14	0.400	0.469	-17.2	105	0.00
32	Benzo(a)anthracene	0.400	0.462	-15.5	113	0.00
33	Chrysene	0.400	0.452	-13.0	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.412	-3.0	111	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.466	-16.5	116	0.00
36 I	Perylene-d12	0.400	0.400	0.0	112	0.00
37	Benzo(b)fluoranthene	0.400	0.403	-0.8	109	0.00
38	Benzo(k)fluoranthene	0.400	0.423	-5.7	110	0.00
39 C	Benzo(a)pyrene	0.400	0.377	5.8	104	0.00
40	Dibenzo(a,h)anthracene	0.400	0.420	-5.0	116	0.00
41	Benzo(g,h,i)perylene	0.400	0.420	-5.0	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

ClientSampleId :

ICVBE032421