

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
 Data File : BE099248.D
 Acq On : 29 Mar 2019 10:50
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 23:04:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	79	-0.01
2	1,4-Dioxane	0.400	0.443	-10.7	100	0.00
3	n-Nitrosodimethylamine	0.400	0.296	26.0#	77	0.00
4 S	2-Fluorophenol	0.400	0.384	4.0	83	-0.01
5 S	Phenol-d6	0.400	0.376	6.0	83	0.00
6	bis(2-Chloroethyl)ether	0.400	0.433	-8.2	97	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	74	0.00
8 S	Nitrobenzene-d5	0.400	0.383	4.3	87	0.00
9	Naphthalene	0.400	0.452	-13.0	94	-0.01
10	Hexachlorobutadiene	0.400	0.465	-16.3	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.466	-16.5	96	0.00
12	2-Methylnaphthalene	0.400	0.445	-11.2	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	77	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.388	3.0	75	0.00
15 S	2-Fluorobiphenyl	0.400	0.457	-14.2	98	0.00
16	Acenaphthylene	0.400	0.406	-1.5	93	0.00
17	Acenaphthene	0.400	0.432	-8.0	96	0.00
18	Fluorene	0.400	0.437	-9.2	98	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	83	-0.01
20	4-Bromophenyl-phenylether	0.400	0.395	1.3	98	-0.01
21	Hexachlorobenzene	0.400	0.416	-4.0	103	-0.01
22	Pentachlorophenol	0.400	0.438	-9.5	109	-0.01
23	Phenanthrene	0.400	0.421	-5.2	103	-0.01
24	Anthracene	0.400	0.393	1.8	101	0.00
25 SURR	Fluoranthene-d10	0.400	0.461	-15.3	112	0.00
26	Fluoranthene	0.400	0.431	-7.7	108	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	83	-0.01
28	Pyrene	0.400	0.452	-13.0	109	0.00
29 S	Terphenyl-d14	0.400	0.455	-13.7	105	0.00
30	Benzo(a)anthracene	0.400	0.366	8.5	91	0.00
31	Chrysene	0.400	0.429	-7.2	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.287	28.3#	85	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.218	45.5#	57	0.00
34 I	Perylene-d12	0.400	0.400	0.0	58	0.00
35	Benzo(b)fluoranthene	0.400	0.361	9.8	64	0.00
36	Benzo(k)fluoranthene	0.400	0.542	-35.5#	93	0.00
37 C	Benzo(a)pyrene	0.400	0.376	6.0	66	0.00
38	Dibenzo(a,h)anthracene	0.400	0.233	41.8#	44	0.00
39	Benzo(g,h,i)perylene	0.400	0.284	29.0#	52	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
Data File : BE099248.D
Acq On : 29 Mar 2019 10:50
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 29 23:04:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				