

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
 Data File : BE099248.D
 Acq On : 29 Mar 2019 10:50
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 23:04:26 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	-0.01
2	1,4-Dioxane	0.527	0.585	-11.0	100	0.00
3	n-Nitrosodimethylamine	0.295	0.218	26.1#	77	0.00
4 S	2-Fluorophenol	1.194	1.147	3.9	83	-0.01
5 S	Phenol-d6	1.825	1.715	6.0	83	0.00
6	bis(2-Chloroethyl)ether	1.435	1.554	-8.3	97	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
8 S	Nitrobenzene-d5	0.327	0.313	4.3	87	0.00
9	Naphthalene	4.717	5.332	-13.0	94	-0.01
10	Hexachlorobutadiene	0.251	0.292	-16.3	99	0.00
11 SURR	2-Methylnaphthalene-d10	0.771	0.899	-16.6	96	0.00
12	2-Methylnaphthalene	0.843	0.939	-11.4	93	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	77	-0.02
14 S	2,4,6-Tribromophenol	0.141	0.090	36.2#	75	0.00
15 S	2-Fluorobiphenyl	1.631	1.863	-14.2	98	0.00
16	Acenaphthylene	2.039	2.071	-1.6	93	0.00
17	Acenaphthene	1.220	1.318	-8.0	96	0.00
18	Fluorene	1.715	1.873	-9.2	98	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	-0.01
20	4-Bromophenyl-phenylether	0.258	0.255	1.2	98	-0.01
21	Hexachlorobenzene	0.249	0.259	-4.0	103	-0.01
22	Pentachlorophenol	0.033	0.018	45.5#	109	-0.01
23	Phenanthrene	1.175	1.235	-5.1	103	-0.01
24	Anthracene	1.061	1.043	1.7	101	0.00
25 SURR	Fluoranthene-d10	4.847	5.584	-15.2	112	0.00
26	Fluoranthene	1.499	1.614	-7.7	108	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	83	-0.01
28	Pyrene	1.329	1.502	-13.0	109	0.00
29 S	Terphenyl-d14	0.870	0.989	-13.7	105	0.00
30	Benzo(a)anthracene	1.320	1.206	8.6	91	0.00
31	Chrysene	1.411	1.513	-7.2	103	0.00
32	Bis(2-ethylhexyl)phthalate	1.574	1.131	28.1#	85	-0.01
33	Indeno(1,2,3-cd)pyrene	1.335	0.728	45.5#	57	0.00
34 I	Perylene-d12	1.000	1.000	0.0	58	0.00
35	Benzo(b)fluoranthene	1.316	1.188	9.7	64	0.00
36	Benzo(k)fluoranthene	1.392	1.885	-35.4#	93	0.00
37 C	Benzo(a)pyrene	1.229	1.155	6.0	66	0.00
38	Dibenzo(a,h)anthracene	1.092	0.635	41.8#	44#	0.00
39	Benzo(g,h,i)perylene	1.095	0.777	29.0#	52	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
Data File : BE099248.D
Acq On : 29 Mar 2019 10:50
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 29 23:04:26 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			