

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
 Data File : BE099263.D
 Acq On : 29 Mar 2019 20:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 23:32:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	-0.01
2	1,4-Dioxane	0.527	0.503	4.6	103	0.00
3	n-Nitrosodimethylamine	0.295	0.321	-8.8	135	-0.01
4 S	2-Fluorophenol	1.194	1.325	-11.0	115	-0.01
5 S	Phenol-d6	1.825	2.039	-11.7	118	0.00
6	bis(2-Chloroethyl)ether	1.435	1.456	-1.5	109	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
8 S	Nitrobenzene-d5	0.327	0.408	-24.8	139	-0.01
9	Naphthalene	4.717	5.043	-6.9	109	-0.01
10	Hexachlorobutadiene	0.251	0.296	-17.9	124	0.00
11 SURR	2-Methylnaphthalene-d10	0.771	0.847	-9.9	112	0.00
12	2-Methylnaphthalene	0.843	0.890	-5.6	108	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	-0.01
14 S	2,4,6-Tribromophenol	0.141	0.200	-41.8#	204#	0.00
15 S	2-Fluorobiphenyl	1.631	1.756	-7.7	113	0.00
16	Acenaphthylene	2.039	2.171	-6.5	119	0.00
17	Acenaphthene	1.220	1.255	-2.9	112	-0.01
18	Fluorene	1.715	1.752	-2.2	112	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	-0.01
20	4-Bromophenyl-phenylether	0.258	0.266	-3.1	115	-0.01
21	Hexachlorobenzene	0.249	0.260	-4.4	116	-0.01
22	Pentachlorophenol	0.033	0.075	-127.3#	505#	-0.01
23	Phenanthrene	1.175	1.213	-3.2	113	-0.01
24	Anthracene	1.061	1.126	-6.1	122	0.00
25 SURR	Fluoranthene-d10	4.847	5.662	-16.8	127	-0.01
26	Fluoranthene	1.499	1.649	-10.0	124	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	103	-0.01
28	Pyrene	1.329	1.383	-4.1	124	0.00
29 S	Terphenyl-d14	0.870	0.942	-8.3	124	0.00
30	Benzo(a)anthracene	1.320	1.487	-12.7	139	0.00
31	Chrysene	1.411	1.470	-4.2	125	0.00
32	Bis(2-ethylhexyl)phthalate	1.574	2.532	-60.9#	236#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.335	1.442	-8.0	140	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.316	1.342	-2.0	127	0.00
36	Benzo(k)fluoranthene	1.392	1.329	4.5	114	0.00
37 C	Benzo(a)pyrene	1.229	1.281	-4.2	129	0.00
38	Dibenzo(a,h)anthracene	1.092	1.140	-4.4	137	-0.02
39	Benzo(g,h,i)perylene	1.095	1.112	-1.6	130	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
Data File : BE099263.D
Acq On : 29 Mar 2019 20:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 29 23:32:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		