

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
 Data File : BE099265.D
 Acq On : 29 Mar 2019 22:46
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 29 23:38:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	-0.01
2	1,4-Dioxane	0.527	0.555	-5.3	106	0.00
3	n-Nitrosodimethylamine	0.295	0.318	-7.8	125	-0.01
4 S	2-Fluorophenol	1.194	1.396	-16.9	114	-0.01
5 S	Phenol-d6	1.825	2.150	-17.8	117	0.00
6	bis(2-Chloroethyl)ether	1.435	1.578	-10.0	111	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	89	0.00
8 S	Nitrobenzene-d5	0.327	0.392	-19.9	131	-0.01
9	Naphthalene	4.717	5.320	-12.8	114	-0.01
10	Hexachlorobutadiene	0.251	0.304	-21.1	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.771	0.881	-14.3	114	-0.01
12	2-Methylnaphthalene	0.843	0.952	-12.9	114	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	92	-0.01
14 S	2,4,6-Tribromophenol	0.141	0.162	-14.9	161#	0.00
15 S	2-Fluorobiphenyl	1.631	1.921	-17.8	120	0.00
16	Acenaphthylene	2.039	2.248	-10.3	120	0.00
17	Acenaphthene	1.220	1.355	-11.1	117	-0.01
18	Fluorene	1.715	1.872	-9.2	116	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	92	-0.01
20	4-Bromophenyl-phenylether	0.258	0.279	-8.1	119	-0.01
21	Hexachlorobenzene	0.249	0.270	-8.4	119	-0.01
22	Pentachlorophenol	0.033	0.031	6.1	206#	-0.01
23	Phenanthrene	1.175	1.235	-5.1	113	-0.01
24	Anthracene	1.061	1.122	-5.7	120	0.00
25 SURR	Fluoranthene-d10	4.847	5.651	-16.6	125	-0.01
26	Fluoranthene	1.499	1.632	-8.9	121	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	-0.01
28	Pyrene	1.329	1.455	-9.5	124	0.00
29 S	Terphenyl-d14	0.870	0.974	-12.0	122	0.00
30	Benzo(a)anthracene	1.320	1.459	-10.5	130	0.00
31	Chrysene	1.411	1.514	-7.3	122	0.00
32	Bis(2-ethylhexyl)phthalate	1.574	2.047	-30.1#	181#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.335	1.230	7.9	114	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	92	0.00
35	Benzo(b)fluoranthene	1.316	1.338	-1.7	114	0.00
36	Benzo(k)fluoranthene	1.392	1.581	-13.6	123	0.00
37 C	Benzo(a)pyrene	1.229	1.329	-8.1	120	0.00
38	Dibenzo(a,h)anthracene	1.092	1.050	3.8	114	-0.02
39	Benzo(g,h,i)perylene	1.095	0.935	14.6	99	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
Data File : BE099265.D
Acq On : 29 Mar 2019 22:46
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 29 23:38:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			