

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
 Data File : BE099268.D
 Acq On : 30 Mar 2019 00:43
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Mar 30 01:23:31 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Mar 26 07:02:06 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	-0.01
2	1,4-Dioxane	0.527	0.482	8.5	105	0.00
3	n-Nitrosodimethylamine	0.295	0.295	0.0	132	-0.01
4 S	2-Fluorophenol	1.194	1.309	-9.6	122	0.00
5 S	Phenol-d6	1.825	2.046	-12.1	126	0.00
6	bis(2-Chloroethyl)ether	1.435	1.528	-6.5	122	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.327	0.413	-26.3#	147	-0.01
9	Naphthalene	4.717	5.203	-10.3	118	-0.01
10	Hexachlorobutadiene	0.251	0.312	-24.3	136	0.00
11 SURR	2-Methylnaphthalene-d10	0.771	0.870	-12.8	120	-0.01
12	2-Methylnaphthalene	0.843	0.923	-9.5	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	-0.01
14 S	2,4,6-Tribromophenol	0.141	0.213	-51.1#	224#	0.00
15 S	2-Fluorobiphenyl	1.631	1.848	-13.3	122	0.00
16	Acenaphthylene	2.039	2.240	-9.9	126	0.00
17	Acenaphthene	1.220	1.319	-8.1	121	-0.01
18	Fluorene	1.715	1.801	-5.0	119	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	-0.01
20	4-Bromophenyl-phenylether	0.258	0.283	-9.7	122	-0.01
21	Hexachlorobenzene	0.249	0.276	-10.8	123	-0.01
22	Pentachlorophenol	0.033	0.093	-181.8#	624#	-0.01
23	Phenanthrene	1.175	1.255	-6.8	117	-0.01
24	Anthracene	1.061	1.176	-10.8	127	0.00
25 SURR	Fluoranthene-d10	4.847	5.934	-22.4	133	0.00
26	Fluoranthene	1.499	1.696	-13.1	128	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	106	-0.01
28	Pyrene	1.329	1.426	-7.3	131	0.00
29 S	Terphenyl-d14	0.870	0.959	-10.2	129	0.00
30	Benzo(a)anthracene	1.320	1.494	-13.2	143	0.00
31	Chrysene	1.411	1.490	-5.6	129	0.00
32	Bis(2-ethylhexyl)phthalate	1.574	2.531	-60.8#	241#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.335	1.483	-11.1	147	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	102	0.00
35	Benzo(b)fluoranthene	1.316	1.383	-5.1	131	0.00
36	Benzo(k)fluoranthene	1.392	1.407	-1.1	122	0.00
37 C	Benzo(a)pyrene	1.229	1.309	-6.5	132	0.00
38	Dibenzo(a,h)anthracene	1.092	1.237	-13.3	150	-0.02
39	Benzo(g,h,i)perylene	1.095	1.214	-10.9	143	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE032919\
Data File : BE099268.D
Acq On : 30 Mar 2019 00:43
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Mar 30 01:23:31 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE032519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Mar 26 07:02:06 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		