

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040821\
 Data File : BE101256.D
 Acq On : 9 Apr 2021 00:11
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Apr 09 01:30:37 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040621.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 07 10:22:45 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	69	0.00
2	1,4-Dioxane	0.400	0.377	5.8	62	0.00
3	n-Nitrosodimethylamine	0.400	0.473	-18.2	83	0.00
4 S	2-Fluorophenol	0.400	0.486	-21.5	84	0.00
5 S	Phenol-d6	0.400	0.503	-25.7#	91	0.00
6	bis(2-Chloroethyl)ether	0.400	0.432	-8.0	74	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	78	0.00
8 S	Nitrobenzene-d5	0.400	0.467	-16.8	95	0.00
9	Naphthalene	0.400	0.401	-0.3	77	0.00
10	Hexachlorobutadiene	0.400	0.392	2.0	75	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.413	-3.2	82	0.00
12	2-Methylnaphthalene	0.400	0.411	-2.7	83	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
14 S	2,4,6-Tribromophenol	0.400	0.426	-6.5	107	0.00
15 S	2-Fluorobiphenyl	0.400	0.377	5.8	81	0.00
16	Acenaphthylene	0.400	0.404	-1.0	97	0.00
17	Acenaphthene	0.400	0.382	4.5	88	0.00
18	Fluorene	0.400	0.406	-1.5	97	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	106	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.209	47.8#	45	0.00
21	4-Bromophenyl-phenylether	0.400	0.391	2.3	102	-0.01
22	Hexachlorobenzene	0.400	0.372	7.0	92	0.00
23	Atrazine	0.400	0.452	-13.0	132	0.00
24	Pentachlorophenol	0.400	0.121	69.8#	45	0.00
25	Phenanthrene	0.400	0.395	1.3	104	0.00
26	Anthracene	0.400	0.401	-0.3	115	0.00
27 SURR	Fluoranthene-d10	0.400	0.431	-7.7	140	0.00
28	Fluoranthene	0.400	0.426	-6.5	138	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	121	0.00
30	Pyrene	0.400	0.424	-6.0	144	0.00
31 S	Terphenyl-d14	0.400	0.435	-8.7	148	0.00
32	Benzo(a)anthracene	0.400	0.430	-7.5	136	0.00
33	Chrysene	0.400	0.399	0.3	119	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.597	-49.2#	185	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.431	-7.7	119	0.00
36 I	Perylene-d12	0.400	0.400	0.0	110	0.00
37	Benzo(b)fluoranthene	0.400	0.370	7.5	104	0.00
38	Benzo(k)fluoranthene	0.400	0.379	5.3	109	0.00
39 C	Benzo(a)pyrene	0.400	0.395	1.3	115	0.00
40	Dibenzo(a,h)anthracene	0.400	0.422	-5.5	118	0.00
41	Benzo(g,h,i)perylene	0.400	0.407	-1.7	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_E

LabSampleId :

SSTDCCC0.4