



Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040821\
 Data File : BE101255.D
 Acq On : 8 Apr 2021 23:36
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-SIM-BE040621.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 07 10:22:45 2021






AutoFind: Scans 2562, 2563, 2564; Background Corrected with Scan 2555

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.9	50678	PASS
68	69	0.00	2	1.4	674	PASS
69	198	0.00	100	40.6	47973	PASS
70	69	0.00	2	0.5	227	PASS
127	198	10	80	43.2	51019	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	118107	PASS
199	198	5	9	6.6	7783	PASS
275	198	10	60	28.9	34109	PASS
365	198	1	100	3.8	4527	PASS
441	198	0.01	100	13.0	15371	PASS
442	442	50	100	100.0	95576	PASS
443	442	15	24	19.3	18432	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040821\
 Data File : BE101255.D
 Acq On : 8 Apr 2021 23:36
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Quant Time: Apr 09 03:15:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 10:54:22 2021
 Response via : Initial Calibration



TIC: BE101255.D\data.ms

(70) Pentachlorophenol (C)

16.848min (-0.013) 2650085.01 ng




response 103986

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.76
264.00	61.60	64.30
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE040821\
 Data File : BE101255.D
 Acq On : 8 Apr 2021 23:36
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Quant Time: Apr 09 03:15:43 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Mar 31 10:54:22 2021
 Response via : Initial Calibration



TIC: BE101255.D\data.ms

(77) Benzidine**19.439min (-0.013) 4330.76 ng****response 915124**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.17
183.00	13.20	11.90
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
4/8/2021	BNA_E	<u>BE101255.D</u>
Compound Name	Response	Retention Time
DDT	461298	20.667
DDD	17487	20.226
DDE	1115	19.727
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
18602	479900	3.88