

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\  
 Data File : BE100021.D  
 Acq On : 14 Jun 2019 13:55  
 Operator : JU/SJ  
 Sample : K3284-09  
 Misc :  
 ALS Vial : 11 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 BPS1-TT-MW306D-20190609

Quant Time: Jun 15 01:40:03 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jun 10 14:57:20 2019  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|--------------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4      | 7.79  | 152  | 871      | 0.40  | ng    | 0.00     |
| 7) Naphthalene-d8              | 10.60 | 136  | 3181     | 0.40  | ng    | 0.00     |
| 13) Acenaphthene-d10           | 14.47 | 164  | 2407     | 0.40  | ng    | 0.00     |
| 19) Phenanthrene-d10           | 17.22 | 188  | 7223     | 0.40  | ng    | 0.00     |
| 27) Chrysene-d12               | 21.40 | 240  | 9170     | 0.40  | ng    | 0.00     |
| 34) Perylene-d12               | 23.93 | 264  | 10484    | 0.40  | ng    | 0.00     |
| System Monitoring Compounds    |       |      |          |       |       |          |
| 4) 2-Fluorophenol              | 5.36  | 112  | 362      | 0.15  | ng    | 0.00     |
| 5) Phenol-d6                   | 6.98  | 99   | 292      | 0.10  | ng    | 0.00     |
| 8) Nitrobenzene-d5             | 8.98  | 82   | 1081     | 0.36  | ng    | 0.01     |
| 11) 2-Methylnaphthalene-d10    | 12.21 | 152  | 1875     | 0.34  | ng    | 0.00     |
| 14) 2,4,6-Tribromophenol       | 15.97 | 330  | 543      | 0.31  | ng    | 0.00     |
| 15) 2-Fluorobiphenyl           | 13.10 | 172  | 4320     | 0.47  | ng    | 0.00     |
| 25) Fluoranthene-d10           | 19.25 | 212  | 35119    | 0.33  | ng    | 0.00     |
| 29) Terphenyl-d14              | 19.84 | 244  | 9339     | 0.57  | ng    | 0.00     |
| Target Compounds               |       |      |          |       |       |          |
| 2) 1,4-Dioxane                 | 3.27  | 88   | 1798     | 1.395 | ng    | 94       |
| 32) Bis(2-ethylhexyl)phthalate | 21.29 | 149  | 2668     | 0.068 | ng    | # 98     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\  
Data File : BE100021.D  
Acq On : 14 Jun 2019 13:55  
Operator : JU/SJ  
Sample : K3284-09  
Misc :  
ALS Vial : 11 Sample Multiplier: 1

Instrument :  
BNA\_E  
ClientSampleId :  
BPS1-TT-MW306D-20190609

Quant Time: Jun 15 01:40:03 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jun 10 14:57:20 2019  
Response via : Initial Calibration

