

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100039.D
 Acq On : 15 Jun 2019 00:46
 Operator : JU/SJ
 Sample : K3315-10
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 TT101D2-20190610

Quant Time: Jun 15 02:58:24 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.78	152	836	0.40	ng	-0.01
7) Naphthalene-d8	10.60	136	2938	0.40	ng	0.00
13) Acenaphthene-d10	14.47	164	2338	0.40	ng	0.00
19) Phenanthrene-d10	17.21	188	6798	0.40	ng	0.00
27) Chrysene-d12	21.40	240	8761	0.40	ng	0.00
34) Perylene-d12	23.93	264	10078	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.37	112	353	0.16	ng	0.01
5) Phenol-d6	6.98	99	252	0.09	ng	0.01
8) Nitrobenzene-d5	8.98	82	1168	0.42	ng	0.01
11) 2-Methylnaphthalene-d10	12.21	152	1922	0.38	ng	0.00
14) 2,4,6-Tribromophenol	15.97	330	513	0.30	ng	0.00
15) 2-Fluorobiphenyl	13.09	172	4571	0.52	ng	-0.01
25) Fluoranthene-d10	19.24	212	34577	0.35	ng	-0.01
29) Terphenyl-d14	19.84	244	10151	0.64	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.27	88	3012	2.434	ng	Qvalue 91
6) bis(2-Chloroethyl)ether	7.23	93	89	0.037	ng	# 70
32) Bis(2-ethylhexyl)phthalate	21.29	149	1430	0.038	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100039.D
Acq On : 15 Jun 2019 00:46
Operator : JU/SJ
Sample : K3315-10
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
TT101D2-20190610

Quant Time: Jun 15 02:58:24 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

