

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100012.D
 Acq On : 14 Jun 2019 8:03
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 00:58:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	119	0.00
2	1,4-Dioxane	0.592	0.580	2.0	124	0.00
3	n-Nitrosodimethylamine	0.341	0.203	40.5#	70	0.03
4 S	2-Fluorophenol	1.080	1.009	6.6	111	0.00
5 S	Phenol-d6	1.391	1.301	6.5	119	0.00
6	bis(2-Chloroethyl)ether	1.136	1.126	0.9	124	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	114	0.00
8 S	Nitrobenzene-d5	0.379	0.367	3.2	117	0.00
9	Naphthalene	4.318	4.483	-3.8	120	-0.01
10	Hexachlorobutadiene	0.329	0.384	-16.7	131	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.726	-4.0	121	0.00
12	2-Methylnaphthalene	0.702	0.713	-1.6	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.296	0.190	35.8#	76	0.00
15 S	2-Fluorobiphenyl	1.517	1.633	-7.6	122	0.00
16	Acenaphthylene	1.990	2.049	-3.0	114	-0.01
17	Acenaphthene	1.084	1.109	-2.3	114	0.00
18	Fluorene	1.627	1.612	0.9	112	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.227	0.243	-7.0	112	0.00
21	Hexachlorobenzene	0.232	0.250	-7.8	110	0.00
22	Pentachlorophenol	0.073	0.042	42.5#	63	0.00
23	Phenanthrene	0.971	0.986	-1.5	105	0.00
24	Anthracene	0.890	0.819	8.0	94	0.00
25 SURR	Fluoranthene-d10	5.828	5.335	8.5	90	0.00
26	Fluoranthene	1.650	1.548	6.2	92	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
28	Pvrene	1.051	1.214	-15.5	91	0.00
29 S	Terphenyl-d14	0.720	0.770	-6.9	86	0.00
30	Benzo(a)anthracene	1.195	1.245	-4.2	84	0.01
31	Chrysene	1.236	1.434	-16.0	90	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.410	17.8	68	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.535	-4.7	83	0.00
34 I	Perylene-d12	1.000	1.000	0.0	82	0.00
35	Benzo(b)fluoranthene	1.121	1.122	-0.1	86	0.00
36	Benzo(k)fluoranthene	1.172	1.265	-7.9	88	0.00
37 C	Benzo(a)pyrene	1.104	1.139	-3.2	87	0.00
38	Dibenzo(a,h)anthracene	1.112	1.029	7.5	79	0.00
39	Benzo(a,h,i)perylene	1.149	1.173	-2.1	86	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100012.D
Acq On : 14 Jun 2019 8:03
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 00:58:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			