

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100025.D
 Acq On : 14 Jun 2019 16:19
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 01:45:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	132	0.00
2	1,4-Dioxane	0.592	0.563	4.9	134	0.00
3	n-Nitrosodimethylamine	0.341	0.251	26.4#	96	0.02
4 S	2-Fluorophenol	1.080	1.126	-4.3	138	0.00
5 S	Phenol-d6	1.391	1.457	-4.7	148	0.00
6	bis(2-Chloroethyl)ether	1.136	1.085	4.5	133	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	131	-0.01
8 S	Nitrobenzene-d5	0.379	0.391	-3.2	144	0.00
9	Naphthalene	4.318	4.473	-3.6	138	-0.01
10	Hexachlorobutadiene	0.329	0.355	-7.9	139	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.736	-5.4	142	-0.01
12	2-Methylnaphthalene	0.702	0.726	-3.4	138	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	126	-0.01
14 S	2,4,6-Tribromophenol	0.296	0.254	14.2	121	-0.01
15 S	2-Fluorobiphenyl	1.517	1.573	-3.7	139	-0.01
16	Acenaphthylene	1.990	2.047	-2.9	135	-0.01
17	Acenaphthene	1.084	1.062	2.0	129	0.00
18	Fluorene	1.627	1.636	-0.6	134	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.227	0.244	-7.5	135	-0.01
21	Hexachlorobenzene	0.232	0.240	-3.4	126	-0.01
22	Pentachlorophenol	0.073	0.088	-20.5	159#	-0.03
23	Phenanthrene	0.971	0.949	2.3	121	0.00
24	Anthracene	0.890	0.865	2.8	119	0.00
25 SURR	Fluoranthene-d10	5.828	5.212	10.6	105	0.00
26	Fluoranthene	1.650	1.473	10.7	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pvrene	1.051	1.227	-16.7	104	0.00
29 S	Terphenyl-d14	0.720	0.822	-14.2	104	0.00
30	Benzo(a)anthracene	1.195	1.385	-15.9	106	0.00
31	Chrysene	1.236	1.323	-7.0	94	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.866	-8.7	102	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.502	-2.5	93	0.00
34 I	Perylene-d12	1.000	1.000	0.0	92	0.00
35	Benzo(b)fluoranthene	1.121	1.111	0.9	95	0.00
36	Benzo(k)fluoranthene	1.172	1.213	-3.5	95	0.00
37 C	Benzo(a)pyrene	1.104	1.127	-2.1	97	0.00
38	Dibenzo(a,h)anthracene	1.112	1.063	4.4	92	0.00
39	Benzo(a,h,i)perylene	1.149	1.083	5.7	89	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100025.D
Acq On : 14 Jun 2019 16:19
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 01:45:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			