

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100027.D
 Acq On : 14 Jun 2019 17:33
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 01:49:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	-0.01
2	1,4-Dioxane	0.592	0.508	14.2	119	0.01
3	n-Nitrosodimethylamine	0.341	0.236	30.8#	89	0.03
4 S	2-Fluorophenol	1.080	1.111	-2.9	134	0.01
5 S	Phenol-d6	1.391	1.422	-2.2	143	0.00
6	bis(2-Chloroethyl)ether	1.136	1.056	7.0	127	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
8 S	Nitrobenzene-d5	0.379	0.374	1.3	133	0.00
9	Naphthalene	4.318	4.510	-4.4	134	-0.01
10	Hexachlorobutadiene	0.329	0.369	-12.2	139	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.730	-4.6	135	0.00
12	2-Methylnaphthalene	0.702	0.724	-3.1	133	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.296	0.236	20.3	109	0.00
15 S	2-Fluorobiphenyl	1.517	1.587	-4.6	136	-0.01
16	Acenaphthylene	1.990	2.032	-2.1	130	-0.01
17	Acenaphthene	1.084	1.069	1.4	126	0.00
18	Fluorene	1.627	1.602	1.5	127	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.227	0.246	-8.4	130	0.00
21	Hexachlorobenzene	0.232	0.247	-6.5	124	-0.01
22	Pentachlorophenol	0.073	0.067	8.2	115	-0.03
23	Phenanthrene	0.971	0.951	2.1	115	0.00
24	Anthracene	0.890	0.838	5.8	110	0.00
25 SURR	Fluoranthene-d10	5.828	5.173	11.2	100	0.00
26	Fluoranthene	1.650	1.471	10.8	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pvrene	1.051	1.222	-16.3	100	0.00
29 S	Terphenyl-d14	0.720	0.803	-11.5	97	0.00
30	Benzo(a)anthracene	1.195	1.276	-6.8	94	0.00
31	Chrysene	1.236	1.366	-10.5	93	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.709	0.4	90	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.436	2.0	85	0.00
34 I	Perylene-d12	1.000	1.000	0.0	88	0.00
35	Benzo(b)fluoranthene	1.121	1.112	0.8	91	0.00
36	Benzo(k)fluoranthene	1.172	1.208	-3.1	90	0.00
37 C	Benzo(a)pyrene	1.104	1.135	-2.8	93	0.00
38	Dibenzo(a,h)anthracene	1.112	0.991	10.9	82	0.00
39	Benzo(a,h,i)perylene	1.149	1.081	5.9	85	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100027.D
Acq On : 14 Jun 2019 17:33
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 01:49:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			