

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100044.D
 Acq On : 15 Jun 2019 3:45
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 05:59:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	152	-0.01
2	1,4-Dioxane	0.400	0.363	9.3	147	0.00
3	n-Nitrosodimethylamine	0.400	0.290	27.5#	110	0.02
4 S	2-Fluorophenol	0.400	0.424	-6.0	162	0.00
5 S	Phenol-d6	0.400	0.398	0.5	163	0.00
6	bis(2-Chloroethyl)ether	0.400	0.380	5.0	152	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	149	-0.01
8 S	Nitrobenzene-d5	0.400	0.426	-6.5	169	0.00
9	Naphthalene	0.400	0.417	-4.2	157	-0.01
10	Hexachlorobutadiene	0.400	0.444	-11.0	163	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.419	-4.7	159	-0.01
12	2-Methylnaphthalene	0.400	0.416	-4.0	157	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	146	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.294	26.5#	119	0.00
15 S	2-Fluorobiphenyl	0.400	0.417	-4.2	161	-0.01
16	Acenaphthylene	0.400	0.406	-1.5	153	-0.01
17	Acenaphthene	0.400	0.403	-0.8	153	0.00
18	Fluorene	0.400	0.391	2.3	150	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	123	0.00
20	4-Bromophenyl-phenylether	0.400	0.453	-13.2	152	-0.01
21	Hexachlorobenzene	0.400	0.435	-8.7	142	-0.01
22	Pentachlorophenol	0.400	0.540	-35.0#	130	-0.03
23	Phenanthrene	0.400	0.419	-4.7	139	0.00
24	Anthracene	0.400	0.408	-2.0	133	0.00
25 SURR	Fluoranthene-d10	0.400	0.363	9.3	115	0.00
26	Fluoranthene	0.400	0.374	6.5	118	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	105	-0.01
28	Pvrene	0.400	0.484	-21.0	119	0.00
29 S	Terphenyl-d14	0.400	0.450	-12.5	113	0.00
30	Benzo(a)anthracene	0.400	0.435	-8.7	110	0.00
31	Chrysene	0.400	0.461	-15.3	111	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.426	-6.5	110	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.424	-6.0	105	0.00
34 I	Perylene-d12	0.400	0.400	0.0	104	0.00
35	Benzo(b)fluoranthene	0.400	0.409	-2.2	111	0.00
36	Benzo(k)fluoranthene	0.400	0.433	-8.2	113	0.00
37 C	Benzo(a)pyrene	0.400	0.417	-4.2	112	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.380	5.0	103	0.00
39	Benzo(a,h,i)perylene	0.400	0.387	3.3	103	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100044.D
Acq On : 15 Jun 2019 3:45
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 05:59:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			