

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100044.D
 Acq On : 15 Jun 2019 3:45
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 05:59:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	152#	-0.01
2	1,4-Dioxane	0.592	0.537	9.3	147	0.00
3	n-Nitrosodimethylamine	0.341	0.248	27.3#	110	0.02
4 S	2-Fluorophenol	1.080	1.145	-6.0	162#	0.00
5 S	Phenol-d6	1.391	1.385	0.4	163#	0.00
6	bis(2-Chloroethyl)ether	1.136	1.080	4.9	152#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	149	-0.01
8 S	Nitrobenzene-d5	0.379	0.403	-6.3	169#	0.00
9	Naphthalene	4.318	4.502	-4.3	157#	-0.01
10	Hexachlorobutadiene	0.329	0.366	-11.2	163#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.731	-4.7	159#	-0.01
12	2-Methylnaphthalene	0.702	0.730	-4.0	157#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	146	-0.01
14 S	2,4,6-Tribromophenol	0.296	0.217	26.7#	119	0.00
15 S	2-Fluorobiphenyl	1.517	1.580	-4.2	161#	-0.01
16	Acenaphthylene	1.990	2.022	-1.6	153#	-0.01
17	Acenaphthene	1.084	1.092	-0.7	153#	0.00
18	Fluorene	1.627	1.590	2.3	150#	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	123	0.00
20	4-Bromophenyl-phenylether	0.227	0.257	-13.2	152#	-0.01
21	Hexachlorobenzene	0.232	0.252	-8.6	142	-0.01
22	Pentachlorophenol	0.073	0.067	8.2	130	-0.03
23	Phenanthrene	0.971	1.018	-4.8	139	0.00
24	Anthracene	0.890	0.907	-1.9	133	0.00
25 SURR	Fluoranthene-d10	5.828	5.293	9.2	115	0.00
26	Fluoranthene	1.650	1.544	6.4	118	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	-0.01
28	Pvrene	1.051	1.271	-20.9	119	0.00
29 S	Terphenyl-d14	0.720	0.811	-12.6	113	0.00
30	Benzo(a)anthracene	1.195	1.301	-8.9	110	0.00
31	Chrysene	1.236	1.426	-15.4	111	-0.01
32	Bis(2-ethylhexyl)phthalate	1.716	1.829	-6.6	110	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.553	-5.9	105	0.00
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.121	1.145	-2.1	111	0.00
36	Benzo(k)fluoranthene	1.172	1.270	-8.4	113	0.00
37 C	Benzo(a)pyrene	1.104	1.151	-4.3	112	-0.01
38	Dibenzo(a,h)anthracene	1.112	1.055	5.1	103	0.00
39	Benzo(a,h,i)perylene	1.149	1.112	3.2	103	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100044.D
Acq On : 15 Jun 2019 3:45
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 05:59:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			