

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100046.D
 Acq On : 15 Jun 2019 6:11
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 06:45:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	148	-0.01
2	1,4-Dioxane	0.400	0.343	14.2	135	0.00
3	n-Nitrosodimethylamine	0.400	0.284	29.0#	104	0.03
4 S	2-Fluorophenol	0.400	0.411	-2.7	152	0.00
5 S	Phenol-d6	0.400	0.418	-4.5	165	0.00
6	bis(2-Chloroethyl)ether	0.400	0.375	6.3	145	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	144	-0.01
8 S	Nitrobenzene-d5	0.400	0.406	-1.5	156	0.00
9	Naphthalene	0.400	0.413	-3.2	151	-0.01
10	Hexachlorobutadiene	0.400	0.453	-13.2	161	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.415	-3.7	153	-0.01
12	2-Methylnaphthalene	0.400	0.411	-2.7	150	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	138	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.302	24.5	116	0.00
15 S	2-Fluorobiphenyl	0.400	0.423	-5.7	155	-0.01
16	Acenaphthylene	0.400	0.404	-1.0	145	-0.01
17	Acenaphthene	0.400	0.402	-0.5	144	0.00
18	Fluorene	0.400	0.399	0.3	145	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	117	0.00
20	4-Bromophenyl-phenylether	0.400	0.454	-13.5	144	-0.01
21	Hexachlorobenzene	0.400	0.456	-14.0	141	-0.01
22	Pentachlorophenol	0.400	0.536	-34.0#	120	-0.01
23	Phenanthrene	0.400	0.422	-5.5	132	0.00
24	Anthracene	0.400	0.408	-2.0	126	0.00
25 SURR	Fluoranthene-d10	0.400	0.373	6.8	111	0.00
26	Fluoranthene	0.400	0.384	4.0	114	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	104	0.00
28	Pvrene	0.400	0.471	-17.7	115	0.00
29 S	Terphenyl-d14	0.400	0.437	-9.2	109	0.00
30	Benzo(a)anthracene	0.400	0.449	-12.2	113	0.00
31	Chrysene	0.400	0.429	-7.2	103	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.393	1.8	101	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.377	5.8	93	0.00
34 I	Perylene-d12	0.400	0.400	0.0	100	0.00
35	Benzo(b)fluoranthene	0.400	0.412	-3.0	107	0.00
36	Benzo(k)fluoranthene	0.400	0.430	-7.5	107	0.00
37 C	Benzo(a)pyrene	0.400	0.415	-3.7	106	0.00
38	Dibenzo(a,h)anthracene	0.400	0.345	13.8	90	0.00
39	Benzo(a,h,i)perylene	0.400	0.373	6.8	95	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100046.D
Acq On : 15 Jun 2019 6:11
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 06:45:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				