

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100046.D
 Acq On : 15 Jun 2019 6:11
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 15 06:45:51 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	148	-0.01
2	1,4-Dioxane	0.592	0.508	14.2	135	0.00
3	n-Nitrosodimethylamine	0.341	0.242	29.0#	104	0.03
4 S	2-Fluorophenol	1.080	1.108	-2.6	152#	0.00
5 S	Phenol-d6	1.391	1.452	-4.4	165#	0.00
6	bis(2-Chloroethyl)ether	1.136	1.064	6.3	145	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	144	-0.01
8 S	Nitrobenzene-d5	0.379	0.385	-1.6	156#	0.00
9	Naphthalene	4.318	4.463	-3.4	151#	-0.01
10	Hexachlorobutadiene	0.329	0.373	-13.4	161#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.724	-3.7	153#	-0.01
12	2-Methylnaphthalene	0.702	0.720	-2.6	150#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	138	-0.01
14 S	2,4,6-Tribromophenol	0.296	0.224	24.3	116	0.00
15 S	2-Fluorobiphenyl	1.517	1.602	-5.6	155#	-0.01
16	Acenaphthylene	1.990	2.012	-1.1	145	-0.01
17	Acenaphthene	1.084	1.090	-0.6	144	0.00
18	Fluorene	1.627	1.622	0.3	145	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
20	4-Bromophenyl-phenylether	0.227	0.257	-13.2	144	-0.01
21	Hexachlorobenzene	0.232	0.264	-13.8	141	-0.01
22	Pentachlorophenol	0.073	0.066	9.6	120	-0.01
23	Phenanthrene	0.971	1.023	-5.4	132	0.00
24	Anthracene	0.890	0.907	-1.9	126	0.00
25 SURR	Fluoranthene-d10	5.828	5.428	6.9	111	0.00
26	Fluoranthene	1.650	1.583	4.1	114	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	1.051	1.239	-17.9	115	0.00
29 S	Terphenyl-d14	0.720	0.787	-9.3	109	0.00
30	Benzo(a)anthracene	1.195	1.342	-12.3	113	0.00
31	Chrysene	1.236	1.326	-7.3	103	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.685	1.8	101	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.383	5.7	93	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.121	1.155	-3.0	107	0.00
36	Benzo(k)fluoranthene	1.172	1.261	-7.6	107	0.00
37 C	Benzo(a)pyrene	1.104	1.145	-3.7	106	0.00
38	Dibenzo(a,h)anthracene	1.112	0.960	13.7	90	0.00
39	Benzo(a,h,i)perylene	1.149	1.070	6.9	95	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100046.D
Acq On : 15 Jun 2019 6:11
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 15 06:45:51 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			