

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100057.D
 Acq On : 15 Jun 2019 13:18
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 17 01:26:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	134	-0.02
2	1,4-Dioxane	0.400	0.353	11.8	126	0.00
3	n-Nitrosodimethylamine	0.400	0.355	11.3	118	0.03
4 S	2-Fluorophenol	0.400	0.391	2.3	131	0.00
5 S	Phenol-d6	0.400	0.414	-3.5	149	0.00
6	bis(2-Chloroethyl)ether	0.400	0.368	8.0	130	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	127	-0.01
8 S	Nitrobenzene-d5	0.400	0.405	-1.3	137	0.00
9	Naphthalene	0.400	0.419	-4.7	135	-0.01
10	Hexachlorobutadiene	0.400	0.459	-14.8	144	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.423	-5.7	138	0.00
12	2-Methylnaphthalene	0.400	0.419	-4.7	136	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	128	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.267	33.3#	95	0.00
15 S	2-Fluorobiphenyl	0.400	0.414	-3.5	140	-0.01
16	Acenaphthylene	0.400	0.396	1.0	131	-0.01
17	Acenaphthene	0.400	0.394	1.5	131	0.00
18	Fluorene	0.400	0.387	3.3	131	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.400	0.417	-4.2	129	-0.01
21	Hexachlorobenzene	0.400	0.424	-6.0	127	-0.01
22	Pentachlorophenol	0.400	0.438	-9.5	68	0.00
23	Phenanthrene	0.400	0.388	3.0	118	0.00
24	Anthracene	0.400	0.366	8.5	110	0.00
25 SURR	Fluoranthene-d10	0.400	0.346	13.5	100	0.00
26	Fluoranthene	0.400	0.355	11.3	103	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	89	0.00
28	Pvrene	0.400	0.479	-19.7	100	0.00
29 S	Terphenyl-d14	0.400	0.445	-11.2	95	0.00
30	Benzo(a)anthracene	0.400	0.440	-10.0	95	0.00
31	Chrysene	0.400	0.450	-12.5	93	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.393	1.8	87	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.369	7.8	78	0.00
34 I	Perylene-d12	0.400	0.400	0.0	85	0.00
35	Benzo(b)fluoranthene	0.400	0.404	-1.0	90	0.00
36	Benzo(k)fluoranthene	0.400	0.431	-7.7	92	0.00
37 C	Benzo(a)pyrene	0.400	0.416	-4.0	91	0.00
38	Dibenzo(a,h)anthracene	0.400	0.327	18.3	73	0.00
39	Benzo(a,h,i)perylene	0.400	0.361	9.8	79	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100057.D
Acq On : 15 Jun 2019 13:18
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 17 01:26:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				