

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
 Data File : BE100057.D
 Acq On : 15 Jun 2019 13:18
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 17 01:26:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 10 14:57:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	134	-0.02
2	1,4-Dioxane	0.592	0.523	11.7	126	0.00
3	n-Nitrosodimethylamine	0.341	0.303	11.1	118	0.03
4 S	2-Fluorophenol	1.080	1.055	2.3	131	0.00
5 S	Phenol-d6	1.391	1.440	-3.5	149	0.00
6	bis(2-Chloroethyl)ether	1.136	1.046	7.9	130	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	127	-0.01
8 S	Nitrobenzene-d5	0.379	0.383	-1.1	137	0.00
9	Naphthalene	4.318	4.525	-4.8	135	-0.01
10	Hexachlorobutadiene	0.329	0.378	-14.9	144	-0.01
11 SURR	2-Methylnaphthalene-d10	0.698	0.738	-5.7	138	0.00
12	2-Methylnaphthalene	0.702	0.735	-4.7	136	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	128	-0.01
14 S	2,4,6-Tribromophenol	0.296	0.197	33.4#	95	0.00
15 S	2-Fluorobiphenyl	1.517	1.569	-3.4	140	-0.01
16	Acenaphthylene	1.990	1.971	1.0	131	-0.01
17	Acenaphthene	1.084	1.067	1.6	131	0.00
18	Fluorene	1.627	1.574	3.3	131	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	113	0.00
20	4-Bromophenyl-phenylether	0.227	0.236	-4.0	129	-0.01
21	Hexachlorobenzene	0.232	0.245	-5.6	127	-0.01
22	Pentachlorophenol	0.073	0.039	46.6#	68	0.00
23	Phenanthrene	0.971	0.941	3.1	118	0.00
24	Anthracene	0.890	0.814	8.5	110	0.00
25 SURR	Fluoranthene-d10	5.828	5.036	13.6	100	0.00
26	Fluoranthene	1.650	1.466	11.2	103	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	89	0.00
28	Pvrene	1.051	1.259	-19.8	100	0.00
29 S	Terphenyl-d14	0.720	0.801	-11.3	95	0.00
30	Benzo(a)anthracene	1.195	1.315	-10.0	95	0.00
31	Chrysene	1.236	1.390	-12.5	93	0.00
32	Bis(2-ethylhexyl)phthalate	1.716	1.687	1.7	87	0.00
33	Indeno(1,2,3-cd)pyrene	1.466	1.354	7.6	78	0.00
34 I	Perylene-d12	1.000	1.000	0.0	85	0.00
35	Benzo(b)fluoranthene	1.121	1.131	-0.9	90	0.00
36	Benzo(k)fluoranthene	1.172	1.264	-7.8	92	0.00
37 C	Benzo(a)pyrene	1.104	1.149	-4.1	91	0.00
38	Dibenzo(a,h)anthracene	1.112	0.908	18.3	73	0.00
39	Benzo(a,h,i)perylene	1.149	1.036	9.8	79	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE061419\
Data File : BE100057.D
Acq On : 15 Jun 2019 13:18
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 17 01:26:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE061019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 10 14:57:20 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			