

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE111324\
 Data File : BE101593.D
 Acq On : 13 Nov 2024 13:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 13 13:41:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE110624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 00:01:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	53	-0.02
2	1,4-Dioxane	40.000	11.616	71.0#	16	-0.02
3	Pyridine	40.000	12.176	69.6#	16	0.02
4	n-Nitrosodimethylamine	40.000	11.628	70.9#	16	-0.02
5 S	2-Fluorophenol	80.000	51.344	35.8#	35	-0.01
6	Aniline	40.000	23.656	40.9#	27	0.00
7 S	Phenol-d6	80.000	79.702	0.4	53	-0.01
8	2-Chlorophenol	40.000	38.101	4.7	51	-0.01
9	Benzaldehyde	40.000	34.111	14.7	44	-0.02
10 C	Phenol	40.000	39.095	2.3	52	0.00
11	bis(2-Chloroethyl)ether	40.000	51.792	-29.5#	82	-0.02
12	1,3-Dichlorobenzene	40.000	40.438	-1.1	55	-0.02
13 C	1,4-Dichlorobenzene	40.000	40.199	-0.5	54	-0.02
14	1,2-Dichlorobenzene	40.000	40.970	-2.4	55	-0.02
15	Benzyl Alcohol	40.000	40.241	-0.6	52	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	49.519	-23.8	66	-0.01
17	2-Methylphenol	40.000	39.521	1.2	51	-0.01
18	Hexachloroethane	40.000	44.380	-11.0	60	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	45.004	-12.5	59	-0.01
20	3+4-Methylphenols	40.000	39.109	2.2	51	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	52	-0.01
22	Acetophenone	40.000	42.784	-7.0	57	-0.02
23 S	Nitrobenzene-d5	80.000	90.958	-13.7	60	-0.01
24	Nitrobenzene	40.000	44.427	-11.1	59	-0.02
25	Isophorone	40.000	45.478	-13.7	60	-0.02
26 C	2-Nitrophenol	40.000	36.916	7.7	49	-0.02
27	2,4-Dimethylphenol	40.000	38.696	3.3	52	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.701	-6.8	57	-0.01
29 C	2,4-Dichlorophenol	40.000	36.657	8.4	49	0.00
30	1,2,4-Trichlorobenzene	40.000	38.897	2.8	53	-0.01
31	Naphthalene	40.000	41.109	-2.8	55	-0.01
32	Benzoic acid	40.000	30.293	24.3	41	-0.02
33	4-Chloroaniline	40.000	38.718	3.2	50	0.00
34 C	Hexachlorobutadiene	40.000	40.235	-0.6	54	-0.01
35	Caprolactam	40.000	38.528	3.7	50	-0.01
36 C	4-Chloro-3-methylphenol	40.000	40.599	-1.5	54	0.00
37	2-Methylnaphthalene	40.000	40.538	-1.3	54	-0.01
38	1-Methylnaphthalene	40.000	40.667	-1.7	55	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	52	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.725	3.2	52	-0.01
41 P	Hexachlorocyclopentadiene	40.000	43.490	-8.7	52	-0.02
42 S	2,4,6-Tribromophenol	80.000	75.040	6.2	50	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.932	5.2	50	0.00
44	2,4,5-Trichlorophenol	40.000	37.135	7.2	49	0.00
45 S	2-Fluorobiphenyl	80.000	84.767	-6.0	56	-0.01
46	1,1'-Biphenyl	40.000	40.848	-2.1	55	0.00
47	2-Chloronaphthalene	40.000	40.285	-0.7	54	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE111324\
 Data File : BE101593.D
 Acq On : 13 Nov 2024 13:09
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 13 13:41:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE110624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 07 00:01:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	45.806	-14.5	60	0.00
49	Acenaphthylene	40.000	41.181	-3.0	55	-0.01
50	Dimethylphthalate	40.000	41.133	-2.8	55	-0.01
51	2,6-Dinitrotoluene	40.000	40.094	-0.2	53	0.00
52 C	Acenaphthene	40.000	41.592	-4.0	56	-0.01
53	3-Nitroaniline	40.000	39.921	0.2	51	0.00
54 P	2,4-Dinitrophenol	40.000	28.782	28.0#	37	0.00
55	Dibenzofuran	40.000	40.535	-1.3	54	0.00
56 P	4-Nitrophenol	40.000	26.751	33.1#	33	0.01
57	2,4-Dinitrotoluene	40.000	41.415	-3.5	54	0.00
58	Fluorene	40.000	42.309	-5.8	56	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.898	5.3	50	0.00
60	Diethylphthalate	40.000	42.925	-7.3	58	0.00
61	4-Chlorophenyl-phenylether	40.000	40.605	-1.5	54	0.00
62	4-Nitroaniline	40.000	40.867	-2.2	52	0.00
63	Azobenzene	40.000	47.078	-17.7	63	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	54	0.00
65	4,6-Dinitro-2-methylphenol	40.000	34.658	13.4	45	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.271	-0.7	55	0.00
67	4-Bromophenyl-phenylether	40.000	38.388	4.0	53	0.00
68	Hexachlorobenzene	40.000	38.484	3.8	53	0.00
69	Atrazine	40.000	38.859	2.9	48	0.00
70 C	Pentachlorophenol	40.000	34.105	14.7	45	0.00
71	Phenanthrene	40.000	41.037	-2.6	57	0.00
72	Anthracene	40.000	41.794	-4.5	57	0.00
73	Carbazole	40.000	41.362	-3.4	57	0.00
74	Di-n-butylphthalate	40.000	43.851	-9.6	60	0.00
75 C	Fluoranthene	40.000	42.891	-7.2	59	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	59	0.00
77	Benzydine	40.000	49.463	-23.7	78	0.00
78	Pyrene	40.000	40.418	-1.0	60	0.00
79 S	Terphenyl-d14	80.000	88.121	-10.2	65	0.00
80	Butylbenzylphthalate	40.000	42.503	-6.3	63	0.00
81	Benzo(a)anthracene	40.000	41.348	-3.4	62	0.00
82	3,3'-Dichlorobenzidine	40.000	39.931	0.2	59	0.00
83	Chrysene	40.000	42.300	-5.7	63	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.957	-7.4	64	0.00
85 c	Di-n-octyl phthalate	40.000	43.674	-9.2	65	0.00
86 I	Perylene-d12	20.000	20.000	0.0	59	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.428	-6.1	64	-0.02
88	Benzo(b)fluoranthene	40.000	39.594	1.0	61	0.00
89	Benzo(k)fluoranthene	40.000	43.511	-8.8	65	0.00
90 C	Benzo(a)pyrene	40.000	41.146	-2.9	62	0.00
91	Dibenzo(a,h)anthracene	40.000	42.592	-6.5	64	-0.02
92	Benzo(g,h,i)perylene	40.000	41.569	-3.9	63	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_E\Data\BE111324\
Data File : BE101593.D
Acq On : 13 Nov 2024 13:09
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC040

Quant Time: Nov 13 13:41:46 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_E\Methods\8270-BE110624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 07 00:01:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0