

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050619\
 Data File : BE099548.D
 Acq On : 6 May 2019 12:27
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 07 08:15:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 07 08:07:29 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	141	0.00
2	1,4-Dioxane	0.550	0.501	8.9	149	0.00
3	n-Nitrosodimethylamine	0.332	0.406	-22.3	195#	-0.02
4 S	2-Fluorophenol	1.106	1.177	-6.4	157#	0.00
5 S	Phenol-d6	1.471	1.611	-9.5	158#	0.00
6	bis(2-Chloroethyl)ether	1.269	1.251	1.4	142	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	139	0.02
8 S	Nitrobenzene-d5	0.375	0.426	-13.6	157#	0.01
9	Naphthalene	4.359	4.403	-1.0	138	0.02
10	Hexachlorobutadiene	0.303	0.325	-7.3	146	0.02
11 SURR	2-Methylnaphthalene-d10	0.721	0.695	3.6	135	0.04
12	2-Methylnaphthalene	0.758	0.743	2.0	137	0.04
13 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.04
14 S	2,4,6-Tribromophenol	0.252	0.312	-23.8	161#	0.04
15 S	2-Fluorobiphenyl	1.526	1.622	-6.3	131	0.03
16	Acenaphthylene	1.957	2.027	-3.6	130	0.03
17	Acenaphthene	1.118	1.127	-0.8	127	0.04
18	Fluorene	1.649	1.627	1.3	123	0.05
19 I	Phenanthrene-d10	1.000	1.000	0.0	118	0.05
20	4-Bromophenyl-phenylether	0.233	0.239	-2.6	119	0.05
21	Hexachlorobenzene	0.222	0.228	-2.7	120	0.05
22	Pentachlorophenol	0.100	0.135	-35.0#	188#	0.05
23	Phenanthrene	1.032	1.050	-1.7	118	0.05
24	Anthracene	0.984	1.002	-1.8	120	0.05
25 SURR	Fluoranthene-d10	5.391	5.537	-2.7	121	0.06
26	Fluoranthene	1.546	1.589	-2.8	120	0.06
27 I	Chrysene-d12	1.000	1.000	0.0	121	0.07
28	Pvrene	1.043	1.177	-12.8	121	0.06
29 S	Terphenyl-d14	0.733	0.820	-11.9	121	0.06
30	Benzo(a)anthracene	1.274	1.452	-14.0	122	0.06
31	Chrysene	1.247	1.441	-15.6	123	0.07
32	Bis(2-ethylhexyl)phthalate	2.216	2.365	-6.7	119	0.06
33	Indeno(1,2,3-cd)pyrene	1.474	1.912	-29.7#	138	0.18
34 I	Perylene-d12	1.000	1.000	0.0	130	0.12
35	Benzo(b)fluoranthene	1.153	1.182	-2.5	131	0.10
36	Benzo(k)fluoranthene	1.112	1.141	-2.6	127	0.10
37 C	Benzo(a)pyrene	1.090	1.144	-5.0	133	0.11
38	Dibenzo(a,h)anthracene	1.115	1.218	-9.2	138	0.19
39	Benzo(a,h,i)perylene	1.118	1.222	-9.3	137	0.20

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050619\
Data File : BE099548.D
Acq On : 6 May 2019 12:27
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 07 08:15:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 07 08:07:29 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			