

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050619\
 Data File : BE099555.D
 Acq On : 6 May 2019 17:50
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: May 07 08:33:22 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE042919.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 07 08:07:29 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	165#	0.01
2	1,4-Dioxane	0.550	0.481	12.5	168#	0.00
3	n-Nitrosodimethylamine	0.332	0.431	-29.8#	242#	-0.01
4 S	2-Fluorophenol	1.106	1.109	-0.3	173#	0.00
5 S	Phenol-d6	1.471	1.608	-9.3	184#	0.01
6	bis(2-Chloroethyl)ether	1.269	1.263	0.5	168#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	172#	0.03
8 S	Nitrobenzene-d5	0.375	0.418	-11.5	191#	0.01
9	Naphthalene	4.359	4.403	-1.0	170#	0.03
10	Hexachlorobutadiene	0.303	0.315	-4.0	174#	0.03
11 SURR	2-Methylnaphthalene-d10	0.721	0.715	0.8	172#	0.03
12	2-Methylnaphthalene	0.758	0.754	0.5	171#	0.04
13 I	Acenaphthene-d10	1.000	1.000	0.0	163#	0.04
14 S	2,4,6-Tribromophenol	0.252	0.307	-21.8	205#	0.04
15 S	2-Fluorobiphenyl	1.526	1.586	-3.9	166#	0.03
16	Acenaphthylene	1.957	2.047	-4.6	170#	0.03
17	Acenaphthene	1.118	1.151	-3.0	168#	0.04
18	Fluorene	1.649	1.678	-1.8	164#	0.05
19 I	Phenanthrene-d10	1.000	1.000	0.0	162#	0.05
20	4-Bromophenyl-phenylether	0.233	0.236	-1.3	162#	0.05
21	Hexachlorobenzene	0.222	0.231	-4.1	167#	0.05
22	Pentachlorophenol	0.100	0.116	-16.0	223#	0.06
23	Phenanthrene	1.032	1.026	0.6	158#	0.05
24	Anthracene	0.984	0.986	-0.2	163#	0.05
25 SURR	Fluoranthene-d10	5.391	5.445	-1.0	163#	0.06
26	Fluoranthene	1.546	1.587	-2.7	164#	0.05
27 I	Chrysene-d12	1.000	1.000	0.0	155#	0.06
28	Pvrene	1.043	1.236	-18.5	162#	0.06
29 S	Terphenyl-d14	0.733	0.860	-17.3	162#	0.06
30	Benzo(a)anthracene	1.274	1.431	-12.3	154#	0.06
31	Chrysene	1.247	1.408	-12.9	154#	0.06
32	Bis(2-ethylhexyl)phthalate	2.216	2.351	-6.1	151#	0.06
33	Indeno(1,2,3-cd)pyrene	1.474	1.691	-14.7	156#	0.17
34 I	Perylene-d12	1.000	1.000	0.0	152#	0.12
35	Benzo(b)fluoranthene	1.153	1.204	-4.4	156#	0.10
36	Benzo(k)fluoranthene	1.112	1.145	-3.0	149	0.10
37 C	Benzo(a)pyrene	1.090	1.148	-5.3	155#	0.11
38	Dibenzo(a,h)anthracene	1.115	1.199	-7.5	158#	0.18
39	Benzo(a,h,i)perylene	1.118	1.193	-6.7	156#	0.20

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE050619\
Data File : BE099555.D
Acq On : 6 May 2019 17:50
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 07 08:33:22 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE042919.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 07 08:07:29 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			