

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE051719\
 Data File : BE099641.D
 Acq On : 17 May 2019 8:25
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 17 08:58:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 19:30:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	138	0.00
2	1,4-Dioxane	0.400	0.337	15.8	115	0.00
3	n-Nitrosodimethylamine	0.400	0.279	30.3#	121	0.01
4 S	2-Fluorophenol	0.400	0.328	18.0	123	0.00
5 S	Phenol-d6	0.400	0.351	12.3	135	0.00
6	bis(2-Chloroethyl)ether	0.400	0.354	11.5	135	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	138	0.00
8 S	Nitrobenzene-d5	0.400	0.338	15.5	133	0.00
9	Naphthalene	0.400	0.368	8.0	142	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	139	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.354	11.5	135	0.00
12	2-Methylnaphthalene	0.400	0.350	12.5	134	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	136	0.00
14 S	2,4,6-Tribromophenol	0.400	0.299	25.3#	113	0.00
15 S	2-Fluorobiphenyl	0.400	0.366	8.5	141	0.01
16	Acenaphthylene	0.400	0.356	11.0	137	0.00
17	Acenaphthene	0.400	0.369	7.8	142	0.00
18	Fluorene	0.400	0.352	12.0	135	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	144	0.00
20	4-Bromophenyl-phenylether	0.400	0.340	15.0	138	0.00
21	Hexachlorobenzene	0.400	0.377	5.8	151	0.00
22	Pentachlorophenol	0.400	0.428	-7.0	117	0.01
23	Phenanthrene	0.400	0.354	11.5	143	0.00
24	Anthracene	0.400	0.328	18.0	136	0.00
25 SURR	Fluoranthene-d10	0.400	0.344	14.0	139	0.00
26	Fluoranthene	0.400	0.351	12.3	141	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	149	0.01
28	Pyrene	0.400	0.383	4.3	144	0.00
29 S	Terphenyl-d14	0.400	0.369	7.8	141	0.00
30	Benzo(a)anthracene	0.400	0.401	-0.3	153	0.00
31	Chrysene	0.400	0.381	4.8	141	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.318	20.5	124	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.416	-4.0	152	0.00
34 I	Perylene-d12	0.400	0.400	0.0	148	0.00
35	Benzo(b)fluoranthene	0.400	0.363	9.3	149	0.00
36	Benzo(k)fluoranthene	0.400	0.367	8.3	152	0.00
37 C	Benzo(a)pyrene	0.400	0.367	8.3	150	0.00
38	Dibenzo(a,h)anthracene	0.400	0.379	5.3	154	0.00
39	Benzo(g,h,i)perylene	0.400	0.377	5.8	151	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE051719\
Data File : BE099641.D
Acq On : 17 May 2019 8:25
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 17 08:58:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 19:30:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				