

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE051719\
 Data File : BE099656.D
 Acq On : 17 May 2019 17:26
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 17 18:28:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 19:30:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	157	0.00
2	1,4-Dioxane	0.400	0.320	20.0	124	0.00
3	n-Nitrosodimethylamine	0.400	0.290	27.5#	142	0.00
4 S	2-Fluorophenol	0.400	0.341	14.8	145	0.00
5 S	Phenol-d6	0.400	0.358	10.5	156	0.00
6	bis(2-Chloroethyl)ether	0.400	0.350	12.5	151	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	159	0.00
8 S	Nitrobenzene-d5	0.400	0.344	14.0	155	0.00
9	Naphthalene	0.400	0.362	9.5	161	0.00
10	Hexachlorobutadiene	0.400	0.355	11.3	154	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	157	0.00
12	2-Methylnaphthalene	0.400	0.360	10.0	158	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	163	0.00
14 S	2,4,6-Tribromophenol	0.400	0.313	21.8	141	-0.01
15 S	2-Fluorobiphenyl	0.400	0.356	11.0	164	0.00
16	Acenaphthylene	0.400	0.352	12.0	162	0.00
17	Acenaphthene	0.400	0.358	10.5	165	-0.01
18	Fluorene	0.400	0.356	11.0	163	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	168	0.00
20	4-Bromophenyl-phenylether	0.400	0.355	11.3	168	0.00
21	Hexachlorobenzene	0.400	0.373	6.8	174	0.00
22	Pentachlorophenol	0.400	0.529	-32.3#	226	0.00
23	Phenanthrene	0.400	0.364	9.0	171	0.00
24	Anthracene	0.400	0.352	12.0	170	0.00
25 SURR	Fluoranthene-d10	0.400	0.360	10.0	170	0.00
26	Fluoranthene	0.400	0.367	8.3	171	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	173	0.00
28	Pyrene	0.400	0.399	0.3	174	0.00
29 S	Terphenyl-d14	0.400	0.385	3.8	171	0.00
30	Benzo(a)anthracene	0.400	0.388	3.0	172	0.00
31	Chrysene	0.400	0.401	-0.3	173	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.337	15.8	153	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.405	-1.3	172	0.00
34 I	Perylene-d12	0.400	0.400	0.0	173	0.00
35	Benzo(b)fluoranthene	0.400	0.360	10.0	173	0.00
36	Benzo(k)fluoranthene	0.400	0.374	6.5	180	0.00
37 C	Benzo(a)pyrene	0.400	0.360	10.0	172	0.00
38	Dibenzo(a,h)anthracene	0.400	0.368	8.0	175	0.00
39	Benzo(g,h,i)perylene	0.400	0.362	9.5	170	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE051719\
Data File : BE099656.D
Acq On : 17 May 2019 17:26
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 17 18:28:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 19:30:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				