

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE051719\
 Data File : BE099656.D
 Acq On : 17 May 2019 17:26
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: May 17 18:28:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051419.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 14 19:30:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	157#	0.00
2	1,4-Dioxane	0.590	0.472	20.0	124	0.00
3	n-Nitrosodimethylamine	0.376	0.273	27.4#	142	0.00
4 S	2-Fluorophenol	1.136	0.968	14.8	145	0.00
5 S	Phenol-d6	1.439	1.288	10.5	156#	0.00
6	bis(2-Chloroethyl)ether	1.211	1.060	12.5	151#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	159#	0.00
8 S	Nitrobenzene-d5	0.387	0.333	14.0	155#	0.00
9	Naphthalene	4.256	3.857	9.4	161#	0.00
10	Hexachlorobutadiene	0.311	0.277	10.9	154#	0.00
11 SURR	2-Methylnaphthalene-d10	0.673	0.603	10.4	157#	0.00
12	2-Methylnaphthalene	0.703	0.633	10.0	158#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	163#	0.00
14 S	2,4,6-Tribromophenol	0.274	0.214	21.9	141	-0.01
15 S	2-Fluorobiphenyl	1.540	1.370	11.0	164#	0.00
16	Acenaphthylene	1.889	1.660	12.1	162#	0.00
17	Acenaphthene	1.066	0.955	10.4	165#	-0.01
18	Fluorene	1.529	1.360	11.1	163#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	168#	0.00
20	4-Bromophenyl-phenylether	0.231	0.205	11.3	168#	0.00
21	Hexachlorobenzene	0.227	0.211	7.0	174#	0.00
22	Pentachlorophenol	0.090	0.084	6.7	226#	0.00
23	Phenanthrene	0.993	0.903	9.1	171#	0.00
24	Anthracene	0.916	0.806	12.0	170#	0.00
25 SURR	Fluoranthene-d10	5.320	4.791	9.9	170#	0.00
26	Fluoranthene	1.502	1.378	8.3	171#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	173#	0.00
28	Pyrene	1.020	1.017	0.3	174#	0.00
29 S	Terphenyl-d14	0.720	0.693	3.8	171#	0.00
30	Benzo(a)anthracene	1.209	1.174	2.9	172#	0.00
31	Chrysene	1.245	1.247	-0.2	173#	-0.01
32	Bis(2-ethylhexyl)phthalate	1.864	1.571	15.7	153#	0.00
33	Indeno(1,2,3-cd)pyrene	1.550	1.569	-1.2	172#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	173#	0.00
35	Benzo(b)fluoranthene	1.118	1.006	10.0	173#	0.00
36	Benzo(k)fluoranthene	1.105	1.032	6.6	180#	0.00
37 C	Benzo(a)pyrene	1.083	0.976	9.9	172#	0.00
38	Dibenzo(a,h)anthracene	1.123	1.034	7.9	175#	0.00
39	Benzo(g,h,i)perylene	1.130	1.023	9.5	170#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE051719\
Data File : BE099656.D
Acq On : 17 May 2019 17:26
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: May 17 18:28:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE051419.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 14 19:30:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		