

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060519\
 Data File : BE099805.D
 Acq On : 5 Jun 2019 14:53
 Operator : JU/SJ
 Sample : SSTCICV0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE060519

Quant Time: Jun 05 15:30:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 05 14:35:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	140	0.01
2	1,4-Dioxane	0.580	0.629	-8.4	178#	0.01
3	n-Nitrosodimethylamine	0.308	0.313	-1.6	175#	0.02
4 S	2-Fluorophenol	0.974	1.090	-11.9	183#	0.02
5 S	Phenol-d6	1.252	1.353	-8.1	170#	0.02
6	bis(2-Chloroethyl)ether	1.097	1.161	-5.8	174#	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	140	0.01
8 S	Nitrobenzene-d5	0.351	0.370	-5.4	167#	0.02
9	Naphthalene	4.271	4.453	-4.3	162#	0.02
10	Hexachlorobutadiene	0.325	0.350	-7.7	167#	0.02
11 SURR	2-Methylnaphthalene-d10	0.687	0.692	-0.7	157#	0.01
12	2-Methylnaphthalene	0.681	0.697	-2.3	163#	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	126	0.01
14 S	2,4,6-Tribromophenol	0.233	0.220	5.6	141	0.00
15 S	2-Fluorobiphenyl	1.487	1.681	-13.0	166#	0.01
16	Acenaphthylene	1.930	2.071	-7.3	154#	0.01
17	Acenaphthene	1.056	1.124	-6.4	152#	0.01
18	Fluorene	1.556	1.617	-3.9	144	0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4-Bromophenyl-phenylether	0.223	0.230	-3.1	139	0.01
21	Hexachlorobenzene	0.245	0.250	-2.0	132	0.01
22	Pentachlorophenol	0.045	0.028	37.8#	99	0.01
23	Phenanthrene	0.955	0.978	-2.4	128	0.00
24	Anthracene	0.863	0.835	3.2	128	0.00
25 SURR	Fluoranthene-d10	5.945	5.226	12.1	112	0.00
26	Fluoranthene	1.673	1.503	10.2	112	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
28	Pvrene	0.966	1.087	-12.5	109	0.00
29 S	Terphenyl-d14	0.656	0.739	-12.7	111	0.00
30	Benzo(a)anthracene	1.105	1.124	-1.7	102	0.01
31	Chrysene	1.221	1.261	-3.3	100	0.00
32	Bis(2-ethylhexyl)phthalate	1.269	1.225	3.5	96	0.00
33	Indeno(1,2,3-cd)pyrene	1.463	1.489	-1.8	99	0.02
34 I	Perylene-d12	1.000	1.000	0.0	87	0.01
35	Benzo(b)fluoranthene	1.107	1.150	-3.9	101	0.00
36	Benzo(k)fluoranthene	1.176	1.224	-4.1	101	0.00
37 C	Benzo(a)pyrene	1.074	1.111	-3.4	100	0.00
38	Dibenzo(a,h)anthracene	1.092	1.082	0.9	98	0.02
39	Benzo(a,h,i)perylene	1.147	1.148	-0.1	99	0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060519\
Data File : BE099805.D
Acq On : 5 Jun 2019 14:53
Operator : JU/SJ
Sample : SSTCICV0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE060519

Quant Time: Jun 05 15:30:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 05 14:35:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			