

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060519\
 Data File : BE099808.D
 Acq On : 5 Jun 2019 17:44
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC040

Quant Time: Jun 06 01:45:13 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 05 14:35:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
2	1,4-Dioxane	0.580	0.496	14.5	105	0.01
3	n-Nitrosodimethylamine	0.308	0.262	14.9	110	0.01
4 S	2-Fluorophenol	0.974	0.826	15.2	104	0.01
5 S	Phenol-d6	1.252	1.076	14.1	101	0.00
6	bis(2-Chloroethyl)ether	1.097	1.009	8.0	113	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.351	0.313	10.8	106	0.00
9	Naphthalene	4.271	3.886	9.0	107	0.00
10	Hexachlorobutadiene	0.325	0.295	9.2	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.687	0.606	11.8	104	0.00
12	2-Methylnaphthalene	0.681	0.570	16.3	100	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.233	0.180	22.7	91	0.00
15 S	2-Fluorobiphenyl	1.487	1.287	13.4	101	0.00
16	Acenaphthylene	1.930	1.721	10.8	102	0.00
17	Acenaphthene	1.056	0.918	13.1	99	0.00
18	Fluorene	1.556	1.355	12.9	95	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
20	4-Bromophenyl-phenylether	0.223	0.194	13.0	100	0.01
21	Hexachlorobenzene	0.245	0.208	15.1	93	0.00
22	Pentachlorophenol	0.045	0.028	37.8#	84	0.04
23	Phenanthrene	0.955	0.824	13.7	92	0.00
24	Anthracene	0.863	0.711	17.6	92	0.00
25 SURR	Fluoranthene-d10	5.945	5.196	12.6	94	0.00
26	Fluoranthene	1.673	1.488	11.1	95	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pvrene	0.966	0.936	3.1	96	0.00
29 S	Terphenyl-d14	0.656	0.629	4.1	96	0.00
30	Benzo(a)anthracene	1.105	1.115	-0.9	103	0.00
31	Chrysene	1.221	1.139	6.7	92	0.00
32	Bis(2-ethylhexyl)phthalate	1.269	1.067	15.9	85	0.00
33	Indeno(1,2,3-cd)pyrene	1.463	1.414	3.3	96	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.107	0.994	10.2	98	0.00
36	Benzo(k)fluoranthene	1.176	1.047	11.0	96	0.00
37 C	Benzo(a)pyrene	1.074	0.955	11.1	96	0.00
38	Dibenzo(a,h)anthracene	1.092	0.977	10.5	99	0.00
39	Benzo(a,h,i)perylene	1.147	0.993	13.4	96	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060519\
Data File : BE099808.D
Acq On : 5 Jun 2019 17:44
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC040

Quant Time: Jun 06 01:45:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 05 14:35:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			