

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060519\
 Data File : BE099812.D
 Acq On : 5 Jun 2019 20:15
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc : MDL-SIM-0.1PPM
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 06 01:41:07 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 05 14:35:58 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.580	0.516	11.0	85	0.00
3	n-Nitrosodimethylamine	0.308	0.203	34.1#	66	0.02
4 S	2-Fluorophenol	0.974	0.795	18.4	78	0.00
5 S	Phenol-d6	1.252	1.026	18.1	75	0.00
6	bis(2-Chloroethyl)ether	1.097	0.992	9.6	87	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	81	0.00
8 S	Nitrobenzene-d5	0.351	0.289	17.7	76	0.01
9	Naphthalene	4.271	3.880	9.2	82	0.00
10	Hexachlorobutadiene	0.325	0.295	9.2	82	0.00
11 SURR	2-Methylnaphthalene-d10	0.687	0.608	11.5	80	0.01
12	2-Methylnaphthalene	0.681	0.574	15.7	78	0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	76	0.01
14 S	2,4,6-Tribromophenol	0.233	0.176	24.5	68	0.00
15 S	2-Fluorobiphenyl	1.487	1.342	9.8	80	0.00
16	Acenaphthylene	1.930	1.773	8.1	79	0.00
17	Acenaphthene	1.056	0.941	10.9	77	0.00
18	Fluorene	1.556	1.439	7.5	77	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	75	0.00
20	4-Bromophenyl-phenylether	0.223	0.175	21.5	72	0.01
21	Hexachlorobenzene	0.245	0.210	14.3	75	0.00
22	Pentachlorophenol	0.045	0.017	62.2#	40#	0.03
23	Phenanthrene	0.955	0.828	13.3	74	0.00
24	Anthracene	0.863	0.686	20.5	72	0.00
25 SURR	Fluoranthene-d10	5.945	5.377	9.6	78	0.00
26	Fluoranthene	1.673	1.540	7.9	79	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
28	Pvrene	0.966	0.888	8.1	79	0.00
29 S	Terphenyl-d14	0.656	0.592	9.8	79	0.00
30	Benzo(a)anthracene	1.105	1.093	1.1	88	0.00
31	Chrysene	1.221	1.166	4.5	82	0.00
32	Bis(2-ethylhexyl)phthalate	1.269	1.040	18.0	72	0.00
33	Indeno(1,2,3-cd)pyrene	1.463	1.546	-5.7	92	0.00
34 I	Perylene-d12	1.000	1.000	0.0	90	0.00
35	Benzo(b)fluoranthene	1.107	1.011	8.7	92	0.00
36	Benzo(k)fluoranthene	1.176	1.020	13.3	87	0.00
37 C	Benzo(a)pyrene	1.074	0.979	8.8	91	0.00
38	Dibenzo(a,h)anthracene	1.092	0.959	12.2	90	0.00
39	Benzo(a,h,i)perylene	1.147	1.020	11.1	91	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE060519\
Data File : BE099812.D
Acq On : 5 Jun 2019 20:15
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc : MDL-SIM-0.1PPM
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 06 01:41:07 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE060519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 05 14:35:58 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			