

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070819\
 Data File : BE100408.D
 Acq On : 8 Jul 2019 12:36
 Operator : JU/SJ
 Sample : PB121046BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB121046BL

Quant Time: Jul 08 15:31:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 08 15:19:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	354	0.40	ng	0.02
7) Naphthalene-d8	10.45	136	1182	0.40	ng	0.03
13) Acenaphthene-d10	14.31	164	817	0.40	ng	0.01
19) Phenanthrene-d10	17.07	188	2443	0.40	ng	0.03
27) Chrysene-d12	21.25	240	3311	0.40	ng	0.01
34) Perylene-d12	23.65	264	5773	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.19	112	303	0.34	ng	0.00
5) Phenol-d6	6.93	99	309	0.26	ng	0.08
8) Nitrobenzene-d5	8.89	82	235	0.20	ng	0.08
11) 2-Methylnaphthalene-d10	12.08	152	781	0.35	ng	0.04
14) 2,4,6-Tribromophenol	15.78	330	2	0.00	ng	-0.03
15) 2-Fluorobiphenyl	12.96	172	1208	0.38	ng	0.03
25) Fluoranthene-d10	19.09	212	13518	0.38	ng	0.01
29) Terphenyl-d14	19.70	244	2615	0.38	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070819\
Data File : BE100408.D
Acq On : 8 Jul 2019 12:36
Operator : JU/SJ
Sample : PB121046BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB121046BL

Quant Time: Jul 08 15:31:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 08 15:19:44 2019
Response via : Initial Calibration

