

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070819\
 Data File : BE100407.D
 Acq On : 8 Jul 2019 12:00
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jul 08 15:27:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 08 15:19:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	56	-0.03
2	1,4-Dioxane	0.621	0.650	-4.7	57	-0.04
3	n-Nitrosodimethylamine	0.289	0.325	-12.5	76	-0.02
4 S	2-Fluorophenol	1.020	0.890	12.7	48#	-0.03
5 S	Phenol-d6	1.342	1.086	19.1	45#	0.01
6	bis(2-Chloroethyl)ether	1.102	0.972	11.8	45#	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	50	-0.03
8 S	Nitrobenzene-d5	0.397	0.312	21.4	40#	-0.01
9	Naphthalene	4.411	4.798	-8.8	54	-0.01
10	Hexachlorobutadiene	0.425	0.554	-30.4#	63	-0.03
11 SURR	2-Methylnaphthalene-d10	0.752	0.796	-5.9	52	-0.01
12	2-Methylnaphthalene	0.774	0.738	4.7	50	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	52	-0.01
14 S	2,4,6-Tribromophenol	0.280	0.262	6.4	50#	-0.01
15 S	2-Fluorobiphenyl	1.559	1.722	-10.5	56	-0.01
16	Acenaphthylene	1.920	2.062	-7.4	54	-0.01
17	Acenaphthene	1.088	1.152	-5.9	52	-0.03
18	Fluorene	1.609	1.882	-17.0	59	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	61	-0.01
20	4-Bromophenyl-phenylether	0.255	0.278	-9.0	63	-0.03
21	Hexachlorobenzene	0.265	0.297	-12.1	65	-0.01
22	Pentachlorophenol	0.080	0.025	68.8#	21#	0.08
23	Phenanthrene	0.969	1.002	-3.4	62	-0.03
24	Anthracene	0.873	0.828	5.2	58	-0.01
25 SURR	Fluoranthene-d10	5.805	6.799	-17.1	70	-0.02
26	Fluoranthene	1.544	1.761	-14.1	67	-0.02
27 I	Chrysene-d12	1.000	1.000	0.0	63	-0.01
28	Pyrene	1.059	1.170	-10.5	69	-0.01
29 S	Terphenyl-d14	0.829	0.867	-4.6	66	-0.01
30	Benzo(a)anthracene	1.315	1.241	5.6	58	0.00
31	Chrysene	1.321	1.371	-3.8	64	-0.01
32	Bis(2-ethylhexyl)phthalate	2.255	1.826	19.0	53	-0.01
33	Indeno(1,2,3-cd)pyrene	1.747	1.815	-3.9	61	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	64	-0.02
35	Benzo(b)fluoranthene	1.084	0.985	9.1	58	-0.02
36	Benzo(k)fluoranthene	1.211	1.443	-19.2	74	-0.02
37 C	Benzo(a)pyrene	1.087	1.068	1.7	61	-0.02
38	Dibenzo(a,h)anthracene	1.135	1.054	7.1	59	-0.04
39	Benzo(g,h,i)perylene	1.166	1.328	-13.9	70	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070819\
Data File : BE100407.D
Acq On : 8 Jul 2019 12:00
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 08 15:27:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 08 15:19:44 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		