

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070819\
 Data File : BE100407.D
 Acq On : 8 Jul 2019 12:00
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 08 15:27:21 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 08 15:19:44 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	56	-0.03
2	1,4-Dioxane	0.400	0.419	-4.7	57	-0.04
3	n-Nitrosodimethylamine	0.400	0.450	-12.5	76	-0.02
4 S	2-Fluorophenol	0.400	0.349	12.8	48	-0.03
5 S	Phenol-d6	0.400	0.324	19.0	45	0.01
6	bis(2-Chloroethyl)ether	0.400	0.353	11.8	45	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	50	-0.03
8 S	Nitrobenzene-d5	0.400	0.314	21.5	40	-0.01
9	Naphthalene	0.400	0.435	-8.7	54	-0.01
10	Hexachlorobutadiene	0.400	0.521	-30.3#	63	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.423	-5.7	52	-0.01
12	2-Methylnaphthalene	0.400	0.381	4.8	50	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	52	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.375	6.3	50	-0.01
15 S	2-Fluorobiphenyl	0.400	0.442	-10.5	56	-0.01
16	Acenaphthylene	0.400	0.430	-7.5	54	-0.01
17	Acenaphthene	0.400	0.423	-5.7	52	-0.03
18	Fluorene	0.400	0.468	-17.0	59	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	61	-0.01
20	4-Bromophenyl-phenylether	0.400	0.435	-8.7	63	-0.03
21	Hexachlorobenzene	0.400	0.449	-12.2	65	-0.01
22	Pentachlorophenol	0.400	0.124	69.0#	21	0.08
23	Phenanthrene	0.400	0.413	-3.2	62	-0.03
24	Anthracene	0.400	0.379	5.3	58	-0.01
25 SURR	Fluoranthene-d10	0.400	0.468	-17.0	70	-0.02
26	Fluoranthene	0.400	0.456	-14.0	67	-0.02
27 I	Chrysene-d12	0.400	0.400	0.0	63	-0.01
28	Pyrene	0.400	0.442	-10.5	69	-0.01
29 S	Terphenyl-d14	0.400	0.418	-4.5	66	-0.01
30	Benzo(a)anthracene	0.400	0.378	5.5	58	0.00
31	Chrysene	0.400	0.415	-3.7	64	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.335	16.3	53	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.416	-4.0	61	-0.03
34 I	Perylene-d12	0.400	0.400	0.0	64	-0.02
35	Benzo(b)fluoranthene	0.400	0.363	9.3	58	-0.02
36	Benzo(k)fluoranthene	0.400	0.477	-19.2	74	-0.02
37 C	Benzo(a)pyrene	0.400	0.393	1.8	61	-0.02
38	Dibenzo(a,h)anthracene	0.400	0.371	7.3	59	-0.04
39	Benzo(g,h,i)perylene	0.400	0.455	-13.7	70	-0.04

Data Path : Z:\SVOASRV\HPCHEM1\BNA_E\DATA\BE070819\
Data File : BE100407.D
Acq On : 8 Jul 2019 12:00
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 08 15:27:21 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 08 15:19:44 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				