

Data Path : Z:\HPCHEM1\BNA_E\Data\BE010215\
 Data File : BE089302.D
 Acq On : 2 Jan 2015 10:33
 Operator : TP/IZ
 Sample : SSTD0.455
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.455

Quant Time: Jan 02 11:15:59 2015
 Quant Title :
 QLast Update : Fri Jan 02 10:52:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)

1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2 SURR	1,4-Dioxane-d8	0.601	0.618	-2.8	124	0.01
3	1,4-Dioxane	0.670	0.676	-0.9	124	0.01
4 I	Naphthalene-d8	1.000	1.000	0.0	131	0.00
5	Naphthalene	0.950	1.021	-7.5	131	0.00
6 SURR	2-Methylnaphthalene-d10	0.459	0.496	-8.1	131	0.00
7	2-Methylnaphthalene	0.580	0.633	-9.1	131	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	131	0.00
9	Acenaphthylene	6.314	7.051	-11.7	135	0.00
10 C	Acenaphthene	4.328	4.618	-6.7	132	0.00
11	Fluorene	1.146	1.215	-6.0	129	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	128	0.00
13	Pentachlorophenol	0.041	0.041	0.0	110	0.00
14	Phenanthrene	4.281	4.587	-7.1	128	-0.01
15	Anthracene	3.824	4.173	-9.1	131	0.00
16 SURR	Fluoranthene-d10	0.693	0.710	-2.5	122	0.00
17 C	Fluoranthene	3.812	3.934	-3.2	122	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
19	Pyrene	5.160	5.746	-11.4	122	0.00
20	Benzo(a)anthracene	0.902	1.026	-13.7	125	0.00
21	Chrysene	1.093	1.172	-7.2	118	0.00
22 I	Perylene-d12	1.000	1.000	0.0	125	0.00
23	Benzo(b)fluoranthene	1.026	0.989	3.6	109	0.00
24	Benzo(k)fluoranthene	1.319	1.410	-6.9	125	0.00
25 C	Benzo(a)pyrene	1.026	1.111	-8.3	123	0.00
26	Indeno(1,2,3-cd)pyrene	2.822	2.545	9.8	103	0.00
27	Dibenzo(a,h)anthracene	0.832	0.774	7.0	109	0.00
28	Benzo(g,h,i)perylene	4.060	3.924	3.3	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0