

Data Path : Z:\HPCHEM1\BNA_E\Data\BE010515\
 Data File : BE089311.D
 Acq On : 5 Jan 2015 12:57
 Operator : TP/IZ
 Sample : SSTD0.457
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.457

Quant Time: Jan 05 13:43:32 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 02 10:52:56 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2 SURR	1,4-Dioxane-d8	0.601	0.618	-2.8	108	0.02
3	1,4-Dioxane	0.670	0.680	-1.5	109	0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
5	Naphthalene	0.950	1.023	-7.7	115	0.00
6 SURR	2-Methylnaphthalene-d10	0.459	0.502	-9.4	116	0.00
7	2-Methylnaphthalene	0.580	0.640	-10.3	116	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
9	Acenaphthylene	6.314	7.351	-16.4	125	0.00
10 C	Acenaphthene	4.328	4.583	-5.9	117	0.00
11	Fluorene	1.146	1.217	-6.2	115	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	112	-0.01
13	Pentachlorophenol	0.041	0.050	-22.0#	120	0.00
14	Phenanthrene	4.281	4.684	-9.4	115	-0.01
15	Anthracene	3.824	4.312	-12.8	119	0.00
16 SURR	Fluoranthene-d10	0.693	0.711	-2.6	107	0.00
17 C	Fluoranthene	3.812	3.991	-4.7	109	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
19	Pyrene	5.160	6.095	-18.1	106	0.00
20	Benzo(a)anthracene	0.902	0.973	-7.9	97	0.00
21	Chrysene	1.093	1.184	-8.3	97	0.00
22 I	Perylene-d12	1.000	1.000	0.0	94	0.00
23	Benzo(b)fluoranthene	1.026	0.938	8.6	78	0.00
24	Benzo(k)fluoranthene	1.319	1.447	-9.7	96	0.00
25 C	Benzo(a)pyrene	1.026	1.076	-4.9	89	0.00
26	Indeno(1,2,3-cd)pyrene	2.822	2.498	11.5	76	0.00
27	Dibenzo(a,h)anthracene	0.832	0.785	5.6	83	0.00
28	Benzo(g,h,i)perylene	4.060	3.995	1.6	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0