

Data Path : Z:\HPCHEM1\BNA_E\Data\BE010515\
 Data File : BE089317.D
 Acq On : 5 Jan 2015 17:01
 Operator : TP/IZ
 Sample : SST0.458
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SST0.458

Quant Time: Jan 06 01:04:09 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM01.2-EPA-SIM-BE123014.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jan 06 00:09:35 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2 SURR	1,4-Dioxane-d8	0.601	0.599	0.3	99	0.02
3	1,4-Dioxane	0.670	0.657	1.9	99	0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
5	Naphthalene	0.950	1.021	-7.5	108	0.00
6 SURR	2-Methylnaphthalene-d10	0.459	0.507	-10.5	111	0.00
7	2-Methylnaphthalene	0.580	0.646	-11.4	111	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	113	0.00
9	Acenaphthylene	6.314	7.233	-14.6	119	0.00
10 C	Acenaphthene	4.328	4.595	-6.2	113	0.00
11	Fluorene	1.146	1.229	-7.2	112	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	109	-0.01
13	Pentachlorophenol	0.041	0.057	-39.0#	131	0.00
14	Phenanthrene	4.281	4.675	-9.2	111	-0.01
15	Anthracene	3.824	4.378	-14.5	116	0.00
16 SURR	Fluoranthene-d10	0.693	0.713	-2.9	104	0.00
17 C	Fluoranthene	3.812	4.029	-5.7	106	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
19	Pyrene	5.160	6.116	-18.5	104	0.00
20	Benzo(a)anthracene	0.902	0.995	-10.3	96	0.00
21	Chrysene	1.093	1.171	-7.1	94	0.00
22 I	Perylene-d12	1.000	1.000	0.0	94	0.00
23	Benzo(b)fluoranthene	1.026	0.973	5.2	81	0.00
24	Benzo(k)fluoranthene	1.319	1.395	-5.8	93	0.00
25 C	Benzo(a)pyrene	1.026	1.076	-4.9	89	0.00
26	Indeno(1,2,3-cd)pyrene	2.822	2.569	9.0	78	0.00
27	Dibenzo(a,h)anthracene	0.832	0.776	6.7	82	-0.01
28	Benzo(g,h,i)perylene	4.060	3.934	3.1	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0