

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010515\
 Data File : BE089319.D
 Acq On : 5 Jan 2015 18:07
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 06 13:21:00 2015
 Quant Method : Z:\HPCHEM1\BNA_E\methods\SIM-BE121914.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 20 00:43:57 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	0.00
2	n-Nitrosodimethylamine	0.653	0.653	0.0	125	0.04
3 S	2-Fluorophenol	1.087	1.046	3.8	116	0.00
4 S	Phenol-d6	1.634	1.663	-1.8	122	0.00
5 C	Phenol	1.956	2.256	-15.3	138	0.00
6	bis(2-Chloroethyl)ether	1.506	1.683	-11.8	136	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	123	0.00
8 S	Nitrobenzene-d5	0.374	0.362	3.2	119	0.00
9	Nitrobenzene	0.388	0.426	-9.8	135	0.00
10	2,4-Dimethylphenol	0.292	0.328	-12.3	137	0.00
11 C	2,4-Dichlorophenol	0.244	0.277	-13.5	138	0.00
12	Hexachlorobutadiene	0.138	0.150	-8.7	133	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	0.135	0.147	-8.9	134	0.00
15 S	2-Fluorobiphenyl	1.417	1.413	0.3	121	0.00
16 P	2,4-Dinitrophenol	0.054	0.065	-20.4	292#	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	128	0.00
18	Hexachlorobenzene	0.241	0.261	-8.3	135	0.00
19 C	Pentachlorophenol	0.057	0.073	-28.1#	205#	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
21	Benzidine	0.629	0.668	-6.2	149	0.00
22 S	Terphenyl-d14	0.893	0.873	2.2	129	0.00
23	Benzo(a)anthracene	4.443	4.667	-5.0	142	0.00
24	3,3'-Dichlorobenzidine	0.357	0.379	-6.2	142	0.00
25	Chrysene	4.680	5.366	-14.7	149	0.00
26	Indeno(1,2,3-cd)pyrene	3.653	3.460	5.3	137	0.00
27 I	Perylene-d12	1.000	1.000	0.0	128	0.00
28	Benzo(b)fluoranthene	0.931	1.024	-10.0	140	0.00
29	Benzo(k)fluoranthene	1.270	1.484	-16.9	147	0.00
30 C	Benzo(a)pyrene	0.915	0.964	-5.4	139	0.00
31	Dibenzo(a,h)anthracene	0.784	0.757	3.4	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1