

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010515\
 Data File : BE089319.D
 Acq On : 5 Jan 2015 18:07
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 06 13:21:00 2015
 Quant Method : Z:\HPCHEM1\BNA_E\methods\SIM-BE121914.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 20 00:43:57 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	126	0.00
2	n-Nitrosodimethylamine	1.000	1.001	-0.1	125	0.04
3 S	2-Fluorophenol	1.000	0.962	3.8	116	0.00
4 S	Phenol-d6	1.000	1.018	-1.8	122	0.00
5 C	Phenol	1.000	1.153	-15.3	138	0.00
6	bis(2-Chloroethyl)ether	1.000	1.117	-11.7	136	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	123	0.00
8 S	Nitrobenzene-d5	1.000	0.966	3.4	119	0.00
9	Nitrobenzene	1.000	1.096	-9.6	135	0.00
10	2,4-Dimethylphenol	1.000	1.124	-12.4	137	0.00
11 C	2,4-Dichlorophenol	1.000	1.133	-13.3	138	0.00
12	Hexachlorobutadiene	1.000	1.088	-8.8	133	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	123	0.00
14 S	2,4,6-Tribromophenol	1.000	1.085	-8.5	134	0.00
15 S	2-Fluorobiphenyl	1.000	0.997	0.3	121	0.00
16 P	2,4-Dinitrophenol	1.000	1.193	-19.3	292	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	128	0.00
18	Hexachlorobenzene	1.000	1.086	-8.6	135	0.00
19 C	Pentachlorophenol	1.000	1.274	-27.4#	205	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	132	0.00
21	Benzidine	1.000	1.062	-6.2	149	0.00
22 S	Terphenyl-d14	1.000	0.978	2.2	129	0.00
23	Benzo(a)anthracene	1.000	1.050	-5.0	142	0.00
24	3,3'-Dichlorobenzidine	1.000	1.060	-6.0	142	0.00
25	Chrysene	1.000	1.147	-14.7	149	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.947	5.3	137	0.00
27 I	Perylene-d12	5.000	5.000	0.0	128	0.00
28	Benzo(b)fluoranthene	1.000	1.099	-9.9	140	0.00
29	Benzo(k)fluoranthene	1.000	1.169	-16.9	147	0.00
30 C	Benzo(a)pyrene	1.000	1.054	-5.4	139	0.00
31	Dibenzo(a,h)anthracene	1.000	0.965	3.5	127	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1