

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE010515\
 Data File : BE089324.D
 Acq On : 5 Jan 2015 20:57
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jan 06 14:12:11 2015
 Quant Method : Z:\HPCHEM1\BNA_E\methods\SIM-BE121914.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 20 00:43:57 2014
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	0.00
2	n-Nitrosodimethylamine	0.653	0.808	-23.7	152#	0.00
3 S	2-Fluorophenol	1.087	1.339	-23.2	146	0.05
4 S	Phenol-d6	1.634	1.766	-8.1	128	0.07
5 C	Phenol	1.956	2.216	-13.3	134	0.07
6	bis(2-Chloroethyl)ether	1.506	1.730	-14.9	138	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
8 S	Nitrobenzene-d5	0.374	0.387	-3.5	125	0.00
9	Nitrobenzene	0.388	0.447	-15.2	138	0.00
10	2,4-Dimethylphenol	0.292	0.349	-19.5	143	0.04
11 C	2,4-Dichlorophenol	0.244	0.292	-19.7	143	0.03
12	Hexachlorobutadiene	0.138	0.150	-8.7	130	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.135	0.151	-11.9	131	0.00
15 S	2-Fluorobiphenyl	1.417	1.447	-2.1	117	0.00
16 P	2,4-Dinitrophenol	0.054	0.041#	24.1	174#	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
18	Hexachlorobenzene	0.241	0.268	-11.2	130	0.00
19 C	Pentachlorophenol	0.057	0.079	-38.6#	209#	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	130	0.00
21	Benzidine	0.629	0.587	6.7	129	0.00
22 S	Terphenyl-d14	0.893	0.897	-0.4	130	0.00
23	Benzo(a)anthracene	4.443	5.299	-19.3	159#	0.00
24	3,3'-Dichlorobenzidine	0.357	0.480	-34.5#	177#	0.00
25	Chrysene	4.680	5.260	-12.4	144	0.00
26	Indeno(1,2,3-cd)pyrene	3.653	5.002	-36.9#	196#	0.00
27 I	Perylene-d12	1.000	1.000	0.0	138	0.00
28	Benzo(b)fluoranthene	0.931	1.241	-33.3#	183#	0.00
29	Benzo(k)fluoranthene	1.270	1.422	-12.0	151#	0.00
30 C	Benzo(a)pyrene	0.915	1.229	-34.3#	190#	0.00
31	Dibenzo(a,h)anthracene	0.784	1.087	-38.6#	195#	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 2